

Surrogate Summary

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PS028900.D	PIBLK-PS028900.D	2,4-DCAA	1	500	474	95		39	175
		2,4-DCAA	2	500	492	98		39	175
I.BLK-PS028988.D	PIBLK-PS028988.D	2,4-DCAA	1	500	512	102		39	175
		2,4-DCAA	2	500	489	98		39	175
PB166382BL	PB166382BL	2,4-DCAA	1	500	445	89		39	175
		2,4-DCAA	2	500	404	81		39	175
PB166382BS	PB166382BS	2,4-DCAA	1	500	510	102		39	175
		2,4-DCAA	2	500	465	93		39	175
PB166318TB	PB166318TB	2,4-DCAA	1	500	462	92		39	175
		2,4-DCAA	2	500	313	63		39	175
Q1206-04	JPP-20.1-012725	2,4-DCAA	1	500	385	77		39	175
		2,4-DCAA	2	500	241	48		39	175
Q1206-04MS	JPP-20.1-012725MS	2,4-DCAA	1	500	424	85		39	175
		2,4-DCAA	2	500	267	53		39	175
Q1206-04MSD	JPP-20.1-012725MSD	2,4-DCAA	1	500	424	85		39	175
		2,4-DCAA	2	500	265	53		39	175
Q1206-08	JPP-16.3-012725	2,4-DCAA	1	500	382	76		39	175
		2,4-DCAA	2	500	263	53		39	175
I.BLK-PS028997.D	PIBLK-PS028997.D	2,4-DCAA	1	500	510	102		39	175
		2,4-DCAA	2	500	499	100		39	175