

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: 8151A

DataFile : PS028994.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	JPP-20.1-012725MS											
Q1206-04MS	2,4-D	50	0	50.4	ug/L	101				65	135	
	2,4,5-TP(Silvex)	50	0	79.2	ug/L	158	*			62	139	