

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL01 Date Analyzed: 01/29/2025

Lab Sample No.: PSTDCCC050 Data File : PL093854.D Time Analyzed: 16:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.060	8.953	9.153	53.100	50.000	6.2
Endrin	6.576	6.472	6.672	53.640	50.000	7.3
gamma-BHC (Lindane)	4.328	4.227	4.427	50.860	50.000	1.7
Heptachlor	4.916	4.814	5.014	52.950	50.000	5.9
Heptachlor epoxide	5.685	5.582	5.782	49.820	50.000	-0.4
Methoxychlor	7.503	7.398	7.598	57.780	50.000	15.6
Tetrachloro-m-xylene	3.539	3.439	3.639	50.560	50.000	1.1

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL01 Date Analyzed: 01/29/2025

Lab Sample No.: PSTDCCC050 Data File : PL093854.D Time Analyzed: 16:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.913	7.810	8.010	53.840	50.000	7.7
Endrin	5.638	5.536	5.736	53.380	50.000	6.8
gamma-BHC (Lindane)	3.607	3.507	3.707	50.260	50.000	0.5
Heptachlor	3.946	3.845	4.045	51.740	50.000	3.5
Heptachlor epoxide	4.728	4.627	4.827	50.190	50.000	0.4
Methoxychlor	6.612	6.509	6.709	55.130	50.000	10.3
Tetrachloro-m-xylene	2.774	2.674	2.874	49.850	50.000	-0.3