

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL02 Date Analyzed: 01/29/2025

Lab Sample No.: PSTDCCC050 Data File : PL093864.D Time Analyzed: 19:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.059	8.953	9.153	49.770	50.000	-0.5
Endrin	6.576	6.472	6.672	49.830	50.000	-0.3
gamma-BHC (Lindane)	4.328	4.227	4.427	46.930	50.000	-6.1
Heptachlor	4.916	4.814	5.014	48.830	50.000	-2.3
Heptachlor epoxide	5.685	5.582	5.782	46.430	50.000	-7.1
Methoxychlor	7.502	7.398	7.598	53.490	50.000	7.0
Tetrachloro-m-xylene	3.539	3.439	3.639	46.770	50.000	-6.5

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL02 Date Analyzed: 01/29/2025

Lab Sample No.: PSTDCCC050 Data File : PL093864.D Time Analyzed: 19:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.913	7.810	8.010	50.680	50.000	1.4
Endrin	5.639	5.536	5.736	49.190	50.000	-1.6
gamma-BHC (Lindane)	3.608	3.507	3.707	46.140	50.000	-7.7
Heptachlor	3.946	3.845	4.045	47.650	50.000	-4.7
Heptachlor epoxide	4.729	4.627	4.827	46.420	50.000	-7.2
Methoxychlor	6.612	6.509	6.709	51.050	50.000	2.1
Tetrachloro-m-xylene	2.775	2.674	2.874	46.240	50.000	-7.5