

### CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 01/30/2025

Lab Sample No.: PSTDCCC050 Data File : PL093876.D Time Analyzed: 09:58

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Decachlorobiphenyl   | 9.055 | 8.953     | 9.153 | 48.590             | 50.000            | -2.8 |
| Endrin               | 6.574 | 6.472     | 6.672 | 48.110             | 50.000            | -3.8 |
| gamma-BHC (Lindane)  | 4.327 | 4.227     | 4.427 | 45.140             | 50.000            | -9.7 |
| Heptachlor           | 4.915 | 4.814     | 5.014 | 47.560             | 50.000            | -4.9 |
| Heptachlor epoxide   | 5.683 | 5.582     | 5.782 | 45.130             | 50.000            | -9.7 |
| Methoxychlor         | 7.500 | 7.398     | 7.598 | 51.550             | 50.000            | 3.1  |
| Tetrachloro-m-xylene | 3.539 | 3.439     | 3.639 | 45.390             | 50.000            | -9.2 |

### CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 01/30/2025

Lab Sample No.: PSTDCCC050 Data File : PL093876.D Time Analyzed: 09:58

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Decachlorobiphenyl   | 7.910 | 7.810     | 8.010 | 49.800             | 50.000            | -0.4 |
| Endrin               | 5.636 | 5.536     | 5.736 | 48.870             | 50.000            | -2.3 |
| gamma-BHC (Lindane)  | 3.607 | 3.507     | 3.707 | 46.010             | 50.000            | -8.0 |
| Heptachlor           | 3.945 | 3.845     | 4.045 | 47.440             | 50.000            | -5.1 |
| Heptachlor epoxide   | 4.727 | 4.627     | 4.827 | 46.460             | 50.000            | -7.1 |
| Methoxychlor         | 6.610 | 6.509     | 6.709 | 49.830             | 50.000            | -0.3 |
| Tetrachloro-m-xylene | 2.774 | 2.674     | 2.874 | 45.470             | 50.000            | -9.1 |