

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL06 Date Analyzed: 02/03/2025

Lab Sample No.: PSTDCCC050 Data File : PL093982.D Time Analyzed: 12:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.061	8.953	9.153	48.630	50.000	-2.7
Endrin	6.577	6.472	6.672	51.580	50.000	3.2
gamma-BHC (Lindane)	4.333	4.227	4.427	55.370	50.000	10.7
Heptachlor	4.921	4.814	5.014	55.190	50.000	10.4
Heptachlor epoxide	5.688	5.582	5.782	48.040	50.000	-3.9
Methoxychlor	7.503	7.398	7.598	57.600	50.000	15.2
Tetrachloro-m-xylene	3.544	3.439	3.639	54.550	50.000	9.1

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL06 Date Analyzed: 02/03/2025

Lab Sample No.: PSTDCCC050 Data File : PL093982.D Time Analyzed: 12:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.913	7.810	8.010	46.730	50.000	-6.5
Endrin	5.638	5.536	5.736	52.920	50.000	5.8
gamma-BHC (Lindane)	3.608	3.507	3.707	52.550	50.000	5.1
Heptachlor	3.946	3.845	4.045	53.790	50.000	7.6
Heptachlor epoxide	4.728	4.627	4.827	53.570	50.000	7.1
Methoxychlor	6.612	6.509	6.709	55.210	50.000	10.4
Tetrachloro-m-xylene	2.774	2.674	2.874	54.490	50.000	9.0