

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL07 Date Analyzed: 02/03/2025

Lab Sample No.: PSTDCCC050 Data File : PL094001.D Time Analyzed: 17:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.055	8.953	9.153	50.120	50.000	0.2
Endrin	6.575	6.472	6.672	53.310	50.000	6.6
gamma-BHC (Lindane)	4.328	4.227	4.427	53.780	50.000	7.6
Heptachlor	4.916	4.814	5.014	54.610	50.000	9.2
Heptachlor epoxide	5.684	5.582	5.782	51.770	50.000	3.5
Methoxychlor	7.501	7.398	7.598	56.300	50.000	12.6
Tetrachloro-m-xylene	3.540	3.439	3.639	52.880	50.000	5.8

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL07 Date Analyzed: 02/03/2025

Lab Sample No.: PSTDCCC050 Data File : PL094001.D Time Analyzed: 17:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.909	7.810	8.010	52.420	50.000	4.8
Endrin	5.636	5.536	5.736	55.090	50.000	10.2
gamma-BHC (Lindane)	3.606	3.507	3.707	53.550	50.000	7.1
Heptachlor	3.944	3.845	4.045	53.700	50.000	7.4
Heptachlor epoxide	4.727	4.627	4.827	54.320	50.000	8.6
Methoxychlor	6.609	6.509	6.709	58.410	50.000	16.8
Tetrachloro-m-xylene	2.773	2.674	2.874	53.550	50.000	7.1