

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No. (PEM): PEM - PL093981.D Date Analyzed: 02/03/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:03

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.056	8.960	9.160	14.510	20.000	-27.5
Tetrachloro-m-xylene	3.540	3.490	3.590	18.660	20.000	-6.7
alpha-BHC	3.996	3.950	4.050	9.460	10.000	-5.4
beta-BHC	4.527	4.480	4.580	9.070	10.000	-9.3
gamma-BHC (Lindane)	4.329	4.280	4.380	9.100	10.000	-9.0
Endrin	6.574	6.500	6.640	37.910	50.000	-24.2
4,4'-DDT	7.025	6.950	7.100	80.590	100.000	-19.4
Methoxychlor	7.502	7.430	7.570	195.740	250.000	-21.7

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No. (PEM): PEM - PL093981.D Date Analyzed: 02/03/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:03

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.911	7.810	8.010	14.500	20.000	-27.5
Tetrachloro-m-xylene	2.775	2.720	2.830	18.090	20.000	-9.6
alpha-BHC	3.277	3.230	3.330	8.770	10.000	-12.3
beta-BHC	3.908	3.860	3.960	9.580	10.000	-4.2
gamma-BHC (Lindane)	3.607	3.560	3.660	7.660	10.000	-23.4
Endrin	5.638	5.570	5.710	37.350	50.000	-25.3
4,4'-DDT	6.035	5.960	6.110	88.620	100.000	-11.4
Methoxychlor	6.610	6.540	6.680	208.200	250.000	-16.7