

Surrogate Summary

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL093725.D	PIBLK-PL093725.D	Decachlorobiphenyl	1	20	22.1	111		43	140
		Tetrachloro-m-xylene	1	20	20.8	104		77	126
		Decachlorobiphenyl	2	20	21.9	109		43	140
		Tetrachloro-m-xylene	2	20	20.5	103		77	126
I.BLK-PL093853.D	PIBLK-PL093853.D	Decachlorobiphenyl	1	20	23.0	115		43	140
		Tetrachloro-m-xylene	1	20	20.0	100		77	126
		Decachlorobiphenyl	2	20	22.0	110		43	140
		Tetrachloro-m-xylene	2	20	18.6	93		77	126
PB166353BL	PB166353BL	Decachlorobiphenyl	1	20	27.2	136		43	140
		Tetrachloro-m-xylene	1	20	23.7	119		77	126
		Decachlorobiphenyl	2	20	26.2	131		43	140
		Tetrachloro-m-xylene	2	20	22.3	111		77	126
PB166318TB	PB166318TB	Decachlorobiphenyl	1	20	26.8	134		43	140
		Tetrachloro-m-xylene	1	20	23.6	118		77	126
		Decachlorobiphenyl	2	20	26.1	130		43	140
		Tetrachloro-m-xylene	2	20	21.9	110		77	126
I.BLK-PL093863.D	PIBLK-PL093863.D	Decachlorobiphenyl	1	20	23.1	115		43	140
		Tetrachloro-m-xylene	1	20	19.8	99		77	126
		Decachlorobiphenyl	2	20	22.1	111		43	140
		Tetrachloro-m-xylene	2	20	18.7	93		77	126
I.BLK-PL093874.D	PIBLK-PL093874.D	Decachlorobiphenyl	1	20	22.5	113		43	140
		Tetrachloro-m-xylene	1	20	19.9	100		77	126
		Decachlorobiphenyl	2	20	21.0	105		43	140
		Tetrachloro-m-xylene	2	20	18.1	91		77	126
PB166353BS	PB166353BS	Decachlorobiphenyl	1	20	23.4	117		43	140
		Tetrachloro-m-xylene	1	20	19.9	99		77	126
		Decachlorobiphenyl	2	20	22.5	113		43	140
		Tetrachloro-m-xylene	2	20	19.1	95		77	126
I.BLK-PL093887.D	PIBLK-PL093887.D	Decachlorobiphenyl	1	20	21.4	107		43	140
		Tetrachloro-m-xylene	1	20	18.4	92		77	126
		Decachlorobiphenyl	2	20	21.2	106		43	140
		Tetrachloro-m-xylene	2	20	17.5	88		77	126
Q1206-08	JPP-16.3-012725	Decachlorobiphenyl	1	20	25.6	128		43	140
		Tetrachloro-m-xylene	1	20	19.9	99		77	126
		Decachlorobiphenyl	2	20	24.8	124		43	140
		Tetrachloro-m-xylene	2	20	19.1	96		77	126
I.BLK-PL093904.D	PIBLK-PL093904.D	Decachlorobiphenyl	1	20	20.2	101		43	140
		Tetrachloro-m-xylene	1	20	19.1	96		77	126
		Decachlorobiphenyl	2	20	16.0	80		43	140
		Tetrachloro-m-xylene	2	20	18.6	93		77	126
I.BLK-PL093980.D	PIBLK-PL093980.D	Decachlorobiphenyl	1	20	21.6	108		43	140

Surrogate Summary

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL093980.D	PIBLK-PL093980.D	Tetrachloro-m-xylene	1	20	22.2	111		77	126
		Decachlorobiphenyl	2	20	20.2	101		43	140
Q1206-04	JPP-20.1-012725	Tetrachloro-m-xylene	2	20	21.8	109		77	126
		Decachlorobiphenyl	1	20	23.4	117		43	140
		Tetrachloro-m-xylene	1	20	21.3	106		77	126
		Decachlorobiphenyl	2	20	22.2	111		43	140
		Tetrachloro-m-xylene	2	20	20.4	102		77	126
		Decachlorobiphenyl	1	20	26.2	131		43	140
Q1206-04MS	JPP-20.1-012725MS	Tetrachloro-m-xylene	1	20	21.8	109		77	126
		Decachlorobiphenyl	2	20	27.6	138		43	140
		Tetrachloro-m-xylene	2	20	21.3	106		77	126
		Decachlorobiphenyl	1	20	26.6	133		43	140
Q1206-04MSD	JPP-20.1-012725MSD	Tetrachloro-m-xylene	1	20	21.3	107		77	126
		Decachlorobiphenyl	2	20	27.5	138		43	140
		Tetrachloro-m-xylene	2	20	21.0	105		77	126
		Decachlorobiphenyl	1	20	28.1	140		43	140
		Tetrachloro-m-xylene	1	20	21.1	106		77	126
		Decachlorobiphenyl	2	20	21.9	109		43	140
I.BLK-PL093999.D	PIBLK-PL093999.D	Tetrachloro-m-xylene	2	20	20.5	102		77	126