

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG No.: Q1206

Instrument ID: BNA_F Calibration Date/Time: 02/05/2025 15:16

Lab File ID: BF141447.D Init. Calib. Date(s): 01/29/2025 01/29/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 17:05

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.303	1.309		0.5	
Benzaldehyde	0.960	1.065		10.9	
Phenol-d6	1.657	1.599		-3.5	
Phenol	1.757	1.692		-3.7	20.0
bis(2-Chloroethyl)ether	1.347	1.250		-7.2	
2-Chlorophenol	1.396	1.375		-1.5	
2-Methylphenol	1.135	1.070		-5.7	
2,2-oxybis(1-Chloropropane)	2.342	2.062		-12.0	
Acetophenone	0.475	0.495		4.2	
3+4-Methylphenols	1.393	1.393		0.0	
n-Nitroso-di-n-propylamine	0.985	0.936	0.050	-5.0	
Nitrobenzene-d5	0.379	0.401		5.8	
Hexachloroethane	0.571	0.587		2.8	
Nitrobenzene	0.393	0.410		4.3	
Isophorone	0.656	0.639		-2.6	
2-Nitrophenol	0.185	0.192		3.8	20.0
2,4-Dimethylphenol	0.236	0.229		-3.0	
bis(2-Chloroethoxy)methane	0.418	0.394		-5.7	
2,4-Dichlorophenol	0.289	0.299		3.5	20.0
Naphthalene	1.057	1.071		1.3	
4-Chloroaniline	0.358	0.310		-13.4	
Hexachlorobutadiene	0.194	0.212		9.3	20.0
Caprolactam	0.094	0.093		-1.1	
4-Chloro-3-methylphenol	0.310	0.324		4.5	20.0
2-Methylnaphthalene	0.672	0.677		0.7	
Hexachlorocyclopentadiene	0.168	0.169	0.050	0.6	
2,4,6-Trichlorophenol	0.372	0.384		3.2	20.0
2-Fluorobiphenyl	1.299	1.332		2.5	
2,4,5-Trichlorophenol	0.405	0.414		2.2	
1,1-Biphenyl	1.560	1.541		-1.2	
2-Chloronaphthalene	1.216	1.206		-0.8	
2-Nitroaniline	0.382	0.407		6.5	
Dimethylphthalate	1.351	1.366		1.1	
Acenaphthylene	1.702	1.722		1.2	
2,6-Dinitrotoluene	0.314	0.321		2.2	
3-Nitroaniline	0.307	0.312		1.6	
Acenaphthene	1.216	1.253		3.0	20.0
2,4-Dinitrophenol	0.146	0.168	0.050	15.1	
4-Nitrophenol	0.225	0.235	0.050	4.4	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG No.: Q1206

Instrument ID: BNA_F Calibration Date/Time: 02/05/2025 15:16

Lab File ID: BF141447.D Init. Calib. Date(s): 01/29/2025 01/29/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 17:05

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.627	1.652		1.5	
2,4-Dinitrotoluene	0.406	0.439		8.1	
Diethylphthalate	1.310	1.335		1.9	
4-Chlorophenyl-phenylether	0.617	0.647		4.9	
Fluorene	1.262	1.315		4.2	
4-Nitroaniline	0.302	0.331		9.6	
4,6-Dinitro-2-methylphenol	0.118	0.139		17.8	
n-Nitrosodiphenylamine	0.647	0.640		-1.1	20.0
2,4,6-Tribromophenol	0.183	0.203		10.9	
4-Bromophenyl-phenylether	0.224	0.230		2.7	
Hexachlorobenzene	0.238	0.245		2.9	
Atrazine	0.216	0.205		-5.1	
Pentachlorophenol	0.139	0.152		9.4	20.0
Phenanthrene	1.075	1.125		4.7	
Anthracene	1.053	1.105		4.9	
Carbazole	0.934	1.005		7.6	
Di-n-butylphthalate	1.132	1.175		3.8	
Fluoranthene	1.062	1.201		13.1	20.0
Pyrene	1.920	1.896		-1.3	
Terphenyl-d14	1.297	1.301		0.3	
Butylbenzylphthalate	0.671	0.684		1.9	
3,3-Dichlorobenzidine	0.439	0.417		-5.0	
Benzo (a) anthracene	1.380	1.336		-3.2	
Chrysene	1.265	1.238		-2.1	
Bis(2-ethylhexyl)phthalate	0.851	0.857		0.7	
Di-n-octyl phthalate	1.321	1.226		-7.2	20.0
Benzo (b) fluoranthene	1.258	1.311		4.2	
Benzo (k) fluoranthene	1.114	1.155		3.7	
Benzo (a) pyrene	1.063	1.108		4.2	20.0
Indeno (1,2,3-cd) pyrene	1.291	1.316		1.9	
Dibenzo (a,h) anthracene	1.039	1.048		0.9	
Benzo (g,h,i) perylene	1.080	1.092		1.1	
1,2,4,5-Tetrachlorobenzene	0.569	0.580		1.9	
1,4-Dioxane	0.573	0.522		-8.9	20.0
2,3,4,6-Tetrachlorophenol	0.319	0.343		7.5	

All other compounds must meet a minimum RRF of 0.010.