

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206

EPA Sample No.: SSTDCCC040 Date Analyzed: 01/30/2025

Lab File ID: BF141301.D Time Analyzed: 08:54

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	146467	6.798	575983	8.08	308835	9.84
UPPER LIMIT	292934	7.298	1151970	8.581	617670	10.339
LOWER LIMIT	73233.5	6.298	287992	7.581	154418	9.339
EPA SAMPLE NO.						
01 PB166335BS	142147	6.80	577653	8.08	314839	9.84
02 WC-5MS	167075	6.80	669898	8.08	361408	9.84
03 WC-5MSD	167424	6.80	670473	8.08	356498	9.84

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.