

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG No.: Q1206

Instrument ID: BNA_F Calibration Date/Time: 01/31/2025 17:19

Lab File ID: BF141336.D Init. Calib. Date(s): 01/29/2025 01/29/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 17:05

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.380	1.326		-3.9	
2-Fluorophenol	1.303	1.218		-6.5	
Phenol-d6	1.657	1.534		-7.4	
1,4-Dichlorobenzene	1.547	1.437		-7.2	20.0
2-Methylphenol	1.135	1.062		-6.4	
3+4-Methylphenols	1.393	1.318		-5.4	
Nitrobenzene-d5	0.379	0.348		-8.2	
Hexachloroethane	0.571	0.536		-6.1	
Nitrobenzene	0.393	0.365		-7.1	
Hexachlorobutadiene	0.194	0.179		-7.7	20.0
2,4,6-Trichlorophenol	0.372	0.365		-1.9	20.0
2-Fluorobiphenyl	1.299	1.148		-11.6	
2,4,5-Trichlorophenol	0.405	0.381		-5.9	
2,4-Dinitrotoluene	0.406	0.427		5.2	
2,4,6-Tribromophenol	0.183	0.177		-3.3	
Hexachlorobenzene	0.238	0.232		-2.5	
Pentachlorophenol	0.139	0.186		33.8	20.0
Terphenyl-d14	1.297	1.177		-9.3	

All other compounds must meet a minimum RRF of 0.010.