

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020525\
 Data File : BF141447.D
 Acq On : 05 Feb 2025 15:16
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 05 15:51:32 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 29 17:41:25 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	60	-0.01
2	1,4-Dioxane	0.573	0.522	8.9	59	-0.04
3	Pyridine	1.380	1.287	6.7	59	-0.05
4	n-Nitrosodimethylamine	0.789	0.863	-9.4	69	-0.06
5 S	2-Fluorophenol	1.303	1.309	-0.5	66	0.00
6	Aniline	1.403	1.296	7.6	58	-0.02
7 S	Phenol-d6	1.657	1.599	3.5	64	0.00
8	2-Chlorophenol	1.396	1.375	1.5	64	-0.01
9	Benzaldehyde	0.960	1.065	-10.9	72	-0.01
10 C	Phenol	1.757	1.692	3.7	62	0.00
11	bis(2-Chloroethyl)ether	1.347	1.250	7.2	61	-0.02
12	1,3-Dichlorobenzene	1.540	1.567	-1.8	66	-0.02
13 C	1,4-Dichlorobenzene	1.547	1.560	-0.8	66	-0.01
14	1,2-Dichlorobenzene	1.448	1.477	-2.0	67	-0.01
15	Benzyl Alcohol	1.133	1.192	-5.2	67	-0.01
16	2,2'-oxybis(1-Chloropropane	2.342	2.062	12.0	57	-0.01
17	2-Methylphenol	1.135	1.070	5.7	62	0.00
18	Hexachloroethane	0.571	0.587	-2.8	67	-0.01
19 P	n-Nitroso-di-n-propylamine	0.985	0.936	5.0	63	-0.02
20	3+4-Methylphenols	1.393	1.393	0.0	65	-0.01
21 I	Naphthalene-d8	1.000	1.000	0.0	59	-0.01
22	Acetophenone	0.475	0.495	-4.2	67	-0.01
23 S	Nitrobenzene-d5	0.379	0.401	-5.8	66	-0.01
24	Nitrobenzene	0.393	0.410	-4.3	65	-0.02
25	Isophorone	0.656	0.639	2.6	61	-0.02
26 C	2-Nitrophenol	0.185	0.192	-3.8	64	-0.01
27	2,4-Dimethylphenol	0.236	0.229	3.0	60	0.00
28	bis(2-Chloroethoxy)methane	0.418	0.394	5.7	59	-0.02
29 C	2,4-Dichlorophenol	0.289	0.299	-3.5	65	0.00
30	1,2,4-Trichlorobenzene	0.330	0.344	-4.2	66	-0.01
31	Naphthalene	1.057	1.071	-1.3	64	-0.01
32	Benzoic acid	0.217	0.229	-5.5	61	-0.03
33	4-Chloroaniline	0.358	0.310	13.4	54	-0.01
34 C	Hexachlorobutadiene	0.194	0.212	-9.3	69	-0.01
35	Caprolactam	0.094	0.093	1.1	61	-0.03
36 C	4-Chloro-3-methylphenol	0.310	0.324	-4.5	66	0.00
37	2-Methylnaphthalene	0.672	0.677	-0.7	64	-0.01
38	1-Methylnaphthalene	0.660	0.661	-0.2	64	-0.01
39 I	Acenaphthene-d10	1.000	1.000	0.0	60	-0.01
40	1,2,4,5-Tetrachlorobenzene	0.569	0.580	-1.9	65	-0.01
41 P	Hexachlorocyclopentadiene	0.168	0.169	-0.6	59	-0.01
42 S	2,4,6-Tribromophenol	0.183	0.203	-10.9	70	-0.01
43 C	2,4,6-Trichlorophenol	0.372	0.384	-3.2	63	0.00
44	2,4,5-Trichlorophenol	0.405	0.414	-2.2	66	0.00
45 S	2-Fluorobiphenyl	1.299	1.332	-2.5	67	-0.01
46	1,1'-Biphenyl	1.560	1.541	1.2	63	-0.01
47	2-Chloronaphthalene	1.216	1.206	0.8	64	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020525\
 Data File : BF141447.D
 Acq On : 05 Feb 2025 15:16
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 05 15:51:32 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 29 17:41:25 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.382	0.407	-6.5	67	0.00
49	Acenaphthylene	1.702	1.722	-1.2	65	-0.01
50	Dimethylphthalate	1.351	1.366	-1.1	66	-0.02
51	2,6-Dinitrotoluene	0.314	0.321	-2.2	65	-0.01
52 C	Acenaphthene	1.216	1.253	-3.0	66	-0.01
53	3-Nitroaniline	0.307	0.312	-1.6	62	-0.01
54 P	2,4-Dinitrophenol	0.146	0.168	-15.1	68	-0.01
55	Dibenzofuran	1.627	1.652	-1.5	65	-0.01
56 P	4-Nitrophenol	0.225	0.235	-4.4	63	0.00
57	2,4-Dinitrotoluene	0.406	0.439	-8.1	68	-0.01
58	Fluorene	1.262	1.315	-4.2	67	-0.01
59	2,3,4,6-Tetrachlorophenol	0.319	0.343	-7.5	67	0.00
60	Diethylphthalate	1.310	1.335	-1.9	65	-0.02
61	4-Chlorophenyl-phenylether	0.617	0.647	-4.9	68	-0.01
62	4-Nitroaniline	0.302	0.331	-9.6	68	-0.02
63	Azobenzene	1.311	1.306	0.4	64	-0.01
64 I	Phenanthrene-d10	1.000	1.000	0.0	64	-0.01
65	4,6-Dinitro-2-methylphenol	0.118	0.139	-17.8	71	-0.01
66 c	n-Nitrosodiphenylamine	0.647	0.640	1.1	65	-0.01
67	4-Bromophenyl-phenylether	0.224	0.230	-2.7	68	-0.01
68	Hexachlorobenzene	0.238	0.245	-2.9	69	-0.01
69	Atrazine	0.216	0.205	5.1	62	-0.02
70 C	Pentachlorophenol	0.139	0.152	-9.4	68	-0.01
71	Phenanthrene	1.075	1.125	-4.7	70	-0.01
72	Anthracene	1.053	1.105	-4.9	70	-0.01
73	Carbazole	0.934	1.005	-7.6	72	0.00
74	Di-n-butylphthalate	1.132	1.175	-3.8	69	-0.01
75 C	Fluoranthene	1.062	1.201	-13.1	76	-0.01
76 I	Chrysene-d12	1.000	1.000	0.0	73	-0.01
77	Benzidine	0.641	0.644	-0.5	64	0.00
78	Pyrene	1.920	1.896	1.3	78	-0.01
79 S	Terphenyl-d14	1.297	1.301	-0.3	80	-0.01
80	Butylbenzylphthalate	0.671	0.684	-1.9	79	-0.01
81	Benzo(a)anthracene	1.380	1.336	3.2	76	-0.01
82	3,3'-Dichlorobenzidine	0.439	0.417	5.0	72	-0.01
83	Chrysene	1.265	1.238	2.1	79	-0.01
84	Bis(2-ethylhexyl)phthalate	0.851	0.857	-0.7	79	-0.01
85 c	Di-n-octyl phthalate	1.321	1.226	7.2	72	-0.02
86 I	Perylene-d12	1.000	1.000	0.0	61	-0.02
87	Indeno(1,2,3-cd)pyrene	1.291	1.316	-1.9	64	-0.03
88	Benzo(b)fluoranthene	1.258	1.311	-4.2	72	-0.02
89	Benzo(k)fluoranthene	1.114	1.155	-3.7	65	-0.02
90 C	Benzo(a)pyrene	1.063	1.108	-4.2	67	-0.02
91	Dibenzo(a,h)anthracene	1.039	1.048	-0.9	63	-0.04
92	Benzo(g,h,i)perylene	1.080	1.092	-1.1	64	-0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020525\
Data File : BF141447.D
Acq On : 05 Feb 2025 15:16
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 05 15:51:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012925.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 29 17:41:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0