

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1205-02MSD	Client Sample ID:	VNJ-236MSD					DataFile:	BF141348.D		
Pyridine	500	0	360	ug/L	72		40	*	10	109	20
1,4-Dichlorobenzene	500	0	360	ug/L	72		3		55	125	20
2-Methylphenol	500	0	370	ug/L	74		15		37	126	20
3+4-Methylphenols	500	0	400	ug/L	80		18		31	127	20
Hexachloroethane	500	0	360	ug/L	72		5		49	110	20
Nitrobenzene	500	0	360	ug/L	72		15		62	112	20
Hexachlorobutadiene	500	0	360	ug/L	72		11		52	125	20
2,4,6-Trichlorophenol	500	0	390	ug/L	78		19		78	112	20
2,4,5-Trichlorophenol	500	0	380	ug/L	76		21	*	71	111	20
2,4-Dinitrotoluene	500	0	370	ug/L	74		20		50	142	20
Hexachlorobenzene	500	0	410	ug/L	82		18		72	115	20
Pentachlorophenol	1000	0	770	ug/L	77		20		25	139	20