

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: RUTW01
 Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG No.: Q1206
 Instrument ID: MSVOA_N Calibration Date(s): 01/14/2025 01/14/2025
 Heated Purge: (Y/N) N Calibration Time(s): 14:56 17:19
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF100 = VN085438.D		RRF050 = VN085439.D		RRF020 = VN085440.D			
		RRF010 = VN085441.D		RRF005 = VN085442.D		RRF001 = VN085443.D			
COMPOUND	RRF100	RRF050	RRF020	RRF010	RRF005	RRF001	RRF	% RSD	
Vinyl Chloride	0.697	0.686	0.727	0.711	0.781	0.819	0.737	7.1	
1,1-Dichloroethene	0.548	0.533	0.556	0.526	0.559	0.497	0.537	4.3	
2-Butanone	0.378	0.390	0.398	0.363	0.387	0.386	0.384	3.1	
Carbon Tetrachloride	0.574	0.530	0.579	0.529	0.565	0.567	0.557	4	
Chloroform	1.197	1.175	1.241	1.169	1.253	1.273	1.218	3.6	
Benzene	1.551	1.449	1.527	1.376	1.474	1.400	1.463	4.7	
1,2-Dichloroethane	0.569	0.547	0.575	0.522	0.574	0.517	0.551	4.8	
Trichloroethene	0.362	0.324	0.352	0.310	0.343	0.352	0.341	5.8	
Tetrachloroethene	0.351	0.322	0.365	0.338	0.346	0.323	0.341	4.9	
Chlorobenzene	1.133	1.076	1.154	1.047	1.110	1.051	1.095	4	
1,2-Dichloroethane-d4	0.774	0.831	0.754	0.762	0.914		0.807	8.3	
Dibromofluoromethane	0.359	0.358	0.335	0.310	0.373		0.347	7.1	
Toluene-d8	1.339	1.267	1.207	1.076	1.274		1.232	8.1	
4-Bromofluorobenzene	0.475	0.449	0.410	0.357	0.417		0.422	10.6	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.