

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: SW8260D

Datafile : VN085552.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0129WBSD01	Vinyl chloride	20	16.3	ug/L	81	2		65	117	20
	1,1-Dichloroethene	20	17.9	ug/L	90	0		74	110	20
	2-Butanone	100	110	ug/L	110	10		65	122	20
	Carbon Tetrachloride	20	18.8	ug/L	94	4		77	113	20
	Chloroform	20	19.4	ug/L	97	1		79	113	20
	Benzene	20	19.3	ug/L	97	0		82	109	20
	1,2-Dichloroethane	20	20.6	ug/L	103	3		80	115	20
	Trichloroethene	20	18.3	ug/L	92	3		77	113	20
	Tetrachloroethene	20	19.1	ug/L	96	1		67	123	20
	Chlorobenzene	20	19.2	ug/L	96	1		82	109	20