

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RUTW01
 Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206
 Lab File ID: VN085548.D Date Analyzed: 01/29/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:08
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	190920	8.22	314840	9.10	285899	11.87
UPPER LIMIT	381840	8.724	629680	9.6	571798	12.365
LOWER LIMIT	95460	7.724	157420	8.6	142950	11.365
EPA SAMPLE NO.						
JPP-20.1-012725	159764	8.22	299409	9.10	275083	11.87
JPP-16.3-012725	165855	8.22	316150	9.10	301938	11.87
VN0129WBL01	180587	8.22	341456	9.10	305749	11.87
VN0129WBS01	209601	8.22	351459	9.10	313028	11.87
VN0129WBSD01	199890	8.22	334279	9.10	300606	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.