

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RUTW01
 Lab Code: CHEM Case No.: Q1206 SAS No.: Q1206 SDG NO.: Q1206
 Lab File ID: VN085548.D Date Analyzed: 01/29/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:08
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	149148	13.788				
UPPER LIMIT	298296	14.288				
LOWER LIMIT	74574	13.288				
EPA SAMPLE NO.						
JPP-20.1-012725	91021	13.79				
JPP-16.3-012725	107837	13.79				
VN0129WBL01	123478	13.79				
VN0129WBS01	156507	13.79				
VN0129WBSD01	149604	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.