

Surrogate Summary

SDG No.: Q1206

Client: RU2 Engineering, LLC

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1206-01	JPP-20.1-012725	1,2-Dichloroethane-d4	50	62.7	125	50	163
		Dibromofluoromethane	50	53.6	107	54	147
		Toluene-d8	50	50.4	101	58	134
		4-Bromofluorobenzene	50	49.6	99	29	146
Q1206-05	JPP-16.3-012725	1,2-Dichloroethane-d4	50	59.0	118	50	163
		Dibromofluoromethane	50	52.2	104	54	147
		Toluene-d8	50	50.1	100	58	134
		4-Bromofluorobenzene	50	48.5	97	29	146
VY0203SBL01	VY0203SBL01	1,2-Dichloroethane-d4	50	50.6	101	50	163
		Dibromofluoromethane	50	49.4	99	54	147
		Toluene-d8	50	47.8	96	58	134
		4-Bromofluorobenzene	50	43.4	87	29	146
VY0203SBS01	VY0203SBS01	1,2-Dichloroethane-d4	50	45.9	92	50	163
		Dibromofluoromethane	50	48.8	98	54	147
		Toluene-d8	50	50.2	100	58	134
		4-Bromofluorobenzene	50	49.7	99	29	146
VY0203SBSD01	VY0203SBSD01	1,2-Dichloroethane-d4	50	49.0	98	50	163
		Dibromofluoromethane	50	48.3	97	54	147
		Toluene-d8	50	50.2	100	58	134
		4-Bromofluorobenzene	50	49.7	99	29	146