

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1207

MATRIX: Solid

METHOD: 1030,9012B,9034,9045D,9095B

	NA	NO	YES
1. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
2. Sample Duplicate Analysis Met QC Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
3. Digestion Holding Time Met		✓	
If not met, list number of days exceeded for each sample:			
The Holding Times were met for all samples except for JPP-16.2-012725 of Corrosivity, for JPP-2.1-012725 of Corrosivity.for JPP-20.2-012725 of Corrosivity.for JPP-4.5-012725 of Corrosivity.for JPP-5.1-012725 of Corrosivity as these samples received out of hold.			

ADDITIONAL COMMENTS:

QA REVIEW

Date