

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client:	<u>RU2 Engineering, LLC</u>	SDG No.:	<u>Q1207</u>
Contract:	<u>RUTW01</u>	Lab Code:	<u>CHEM</u>
ICS Source:	<u>EPA</u>	Case No.:	<u>Q1207</u>
		SAS No.:	<u>Q1207</u>
		Instrument ID:	<u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	235000	255000	92	216000	294000	01/30/2025	12:10	LB134493
	Antimony	-3.14			-50	50	01/30/2025	12:10	LB134493
	Arsenic	-3.20			-20	20	01/30/2025	12:10	LB134493
	Barium	1.95	6.0	32	-94	106	01/30/2025	12:10	LB134493
	Beryllium	1.36			-6	6	01/30/2025	12:10	LB134493
	Cadmium	-0.88	1.0	88	-5	7	01/30/2025	12:10	LB134493
	Calcium	218000	245000	89	208000	282000	01/30/2025	12:10	LB134493
	Chromium	56.0	52.0	108	42	62	01/30/2025	12:10	LB134493
	Cobalt	0.92			-30	30	01/30/2025	12:10	LB134493
	Copper	21.7	2.0	1085	-18	22	01/30/2025	12:10	LB134493
	Iron	92700	101000	92	85600	116500	01/30/2025	12:10	LB134493
	Lead	6.67			-12	12	01/30/2025	12:10	LB134493
	Magnesium	234000	255000	92	216000	294000	01/30/2025	12:10	LB134493
	Manganese	3.50	7.0	50	-13	27	01/30/2025	12:10	LB134493
	Nickel	2.07	2.0	104	-38	42	01/30/2025	12:10	LB134493
	Potassium	133			0	0	01/30/2025	12:10	LB134493
	Selenium	-14.4			-20	20	01/30/2025	12:10	LB134493
	Silver	-0.54			-10	10	01/30/2025	12:10	LB134493
	Sodium	23.8			0	0	01/30/2025	12:10	LB134493
	Thallium	11.1			-40	40	01/30/2025	12:10	LB134493
	Vanadium	5.11			-40	40	01/30/2025	12:10	LB134493
	Zinc	1.05			-40	40	01/30/2025	12:10	LB134493
ICSAB01	Aluminum	243000	247000	98	209000	285000	01/30/2025	12:28	LB134493
	Antimony	644	618	104	525	711	01/30/2025	12:28	LB134493
	Arsenic	118	104	114	88.4	120	01/30/2025	12:28	LB134493
	Barium	456	537	85	437	637	01/30/2025	12:28	LB134493
	Beryllium	475	495	96	420	570	01/30/2025	12:28	LB134493
	Cadmium	1020	972	105	826	1120	01/30/2025	12:28	LB134493
	Calcium	224000	235000	95	199000	271000	01/30/2025	12:28	LB134493
	Chromium	583	542	108	460	624	01/30/2025	12:28	LB134493
	Cobalt	514	476	108	404	548	01/30/2025	12:28	LB134493
	Copper	526	511	103	434	588	01/30/2025	12:28	LB134493
	Iron	99900	99300	101	84400	114500	01/30/2025	12:28	LB134493
	Lead	57.1	49.0	116	37	61	01/30/2025	12:28	LB134493
	Magnesium	241000	248000	97	210000	286000	01/30/2025	12:28	LB134493
	Manganese	448	507	88	430	584	01/30/2025	12:28	LB134493
	Nickel	1020	954	107	810	1100	01/30/2025	12:28	LB134493
	Potassium	109			0	0	01/30/2025	12:28	LB134493
	Selenium	36.8	46.0	80	26	66	01/30/2025	12:28	LB134493
	Silver	222	201	110	170	232	01/30/2025	12:28	LB134493
	Sodium	7.13			0	0	01/30/2025	12:28	LB134493
	Thallium	104	108	96	68	148	01/30/2025	12:28	LB134493

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: RU2 Engineering, LLC **SDG No.:** Q1207
Contract: RUTW01 **Lab Code:** CHEM **Case No.:** Q1207 **SAS No.:** Q1207
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	463	491	94	417	565	01/30/2025	12:28	LB134493
	Zinc	1080	952	113	809	1095	01/30/2025	12:28	LB134493