



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: RU2 Engineering, LLC **SDG No.:** Q1207
Contract: RUTW01 **Lab Code:** CHEM **Case No.:** Q1207 **SAS No.:** Q1207
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.19	0.2	97	40 - 160	CV	01/30/2025	09:59	LB134479
CRI01	Arsenic	21.5	20.0	107	40 - 160	P	01/30/2025	12:05	LB134493
	Barium	90.1	100	90	40 - 160	P	01/30/2025	12:05	LB134493
	Cadmium	5.56	6.0	93	40 - 160	P	01/30/2025	12:05	LB134493
	Chromium	9.97	10.0	100	40 - 160	P	01/30/2025	12:05	LB134493
	Lead	10.9	12.0	91	40 - 160	P	01/30/2025	12:05	LB134493
	Selenium	18.4	20.0	92	40 - 160	P	01/30/2025	12:05	LB134493
	Silver	11.1	10.0	111	40 - 160	P	01/30/2025	12:05	LB134493