



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	RU2 Engineering, LLC				SDG No.:	Q1207			
Contract:	RUTW01	Lab Code:	CHEM		Case No.:	Q1207	SAS No.:	Q1207	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB75	Mercury	0.20	+/-0.20	U	0.20	CV	01/30/2025	09:50	LB134479

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: RU2 Engineering, LLC

SDG No.: Q1207

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1207

SAS No.: Q1207

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB87	Mercury	0.20	+/-0.20	U	0.20	CV	01/30/2025	09:54	LB134479
CCB88	Mercury	0.20	+/-0.20	U	0.20	CV	01/30/2025	10:30	LB134479
CCB89	Mercury	0.20	+/-0.20	U	0.20	CV	01/30/2025	10:58	LB134479
CCB90	Mercury	0.20	+/-0.20	U	0.20	CV	01/30/2025	11:17	LB134479

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: RU2 Engineering, LLC

SDG No.: Q1207

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1207

SAS No.: Q1207

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	01/30/2025	12:00	LB134493
	Barium	100	+/-100	U	100	P	01/30/2025	12:00	LB134493
	Cadmium	6.00	+/-6.00	U	6.00	P	01/30/2025	12:00	LB134493
	Chromium	10.0	+/-10.0	U	10.0	P	01/30/2025	12:00	LB134493
	Lead	12.0	+/-12.0	U	12.0	P	01/30/2025	12:00	LB134493
	Selenium	20.0	+/-20.0	U	20.0	P	01/30/2025	12:00	LB134493
	Silver	10.0	+/-10.0	U	10.0	P	01/30/2025	12:00	LB134493

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: RU2 Engineering, LLC

SDG No.: Q1207

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1207

SAS No.: Q1207

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	01/30/2025	12:47	LB134493
	Barium	100	+/-100	U	100	P	01/30/2025	12:47	LB134493
	Cadmium	6.00	+/-6.00	U	6.00	P	01/30/2025	12:47	LB134493
	Chromium	10.0	+/-10.0	U	10.0	P	01/30/2025	12:47	LB134493
	Lead	12.0	+/-12.0	U	12.0	P	01/30/2025	12:47	LB134493
	Selenium	20.0	+/-20.0	U	20.0	P	01/30/2025	12:47	LB134493
	Silver	10.0	+/-10.0	U	10.0	P	01/30/2025	12:47	LB134493
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	01/30/2025	13:37	LB134493
	Barium	100	+/-100	U	100	P	01/30/2025	13:37	LB134493
	Cadmium	6.00	+/-6.00	U	6.00	P	01/30/2025	13:37	LB134493
	Chromium	10.0	+/-10.0	U	10.0	P	01/30/2025	13:37	LB134493
	Lead	12.0	+/-12.0	U	12.0	P	01/30/2025	13:37	LB134493
	Selenium	20.0	+/-20.0	U	20.0	P	01/30/2025	13:37	LB134493
	Silver	10.0	+/-10.0	U	10.0	P	01/30/2025	13:37	LB134493
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	01/30/2025	14:38	LB134493
	Barium	100	+/-100	U	100	P	01/30/2025	14:38	LB134493
	Cadmium	6.00	+/-6.00	U	6.00	P	01/30/2025	14:38	LB134493
	Chromium	10.0	+/-10.0	U	10.0	P	01/30/2025	14:38	LB134493
	Lead	12.0	+/-12.0	U	12.0	P	01/30/2025	14:38	LB134493
	Selenium	20.0	+/-20.0	U	20.0	P	01/30/2025	14:38	LB134493
	Silver	10.0	+/-10.0	U	10.0	P	01/30/2025	14:38	LB134493
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	01/30/2025	15:29	LB134493
	Barium	100	+/-100	U	100	P	01/30/2025	15:29	LB134493
	Cadmium	6.00	+/-6.00	U	6.00	P	01/30/2025	15:29	LB134493
	Chromium	10.0	+/-10.0	U	10.0	P	01/30/2025	15:29	LB134493
	Lead	12.0	+/-12.0	U	12.0	P	01/30/2025	15:29	LB134493
	Selenium	20.0	+/-20.0	U	20.0	P	01/30/2025	15:29	LB134493
	Silver	10.0	+/-10.0	U	10.0	P	01/30/2025	15:29	LB134493
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	01/30/2025	16:28	LB134493
	Barium	100	+/-100	U	100	P	01/30/2025	16:28	LB134493
	Cadmium	6.00	+/-6.00	U	6.00	P	01/30/2025	16:28	LB134493
	Chromium	10.0	+/-10.0	U	10.0	P	01/30/2025	16:28	LB134493
	Lead	12.0	+/-12.0	U	12.0	P	01/30/2025	16:28	LB134493
	Selenium	20.0	+/-20.0	U	20.0	P	01/30/2025	16:28	LB134493
	Silver	10.0	+/-10.0	U	10.0	P	01/30/2025	16:28	LB134493
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	01/30/2025	17:24	LB134493
	Barium	100	+/-100	U	100	P	01/30/2025	17:24	LB134493
	Cadmium	6.00	+/-6.00	U	6.00	P	01/30/2025	17:24	LB134493
	Chromium	10.0	+/-10.0	U	10.0	P	01/30/2025	17:24	LB134493

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>RU2 Engineering, LLC</u>		SDG No.: <u>Q1207</u>							
Contract: <u>RUTW01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1207</u>	SAS No.: <u>Q1207</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	12.0	+/-12.0	U	12.0	P	01/30/2025	17:24	LB134493
	Selenium	20.0	+/-20.0	U	20.0	P	01/30/2025	17:24	LB134493
	Silver	10.0	+/-10.0	U	10.0	P	01/30/2025	17:24	LB134493
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	01/30/2025	18:15	LB134493
	Barium	100	+/-100	U	100	P	01/30/2025	18:15	LB134493
	Cadmium	6.00	+/-6.00	U	6.00	P	01/30/2025	18:15	LB134493
	Chromium	10.0	+/-10.0	U	10.0	P	01/30/2025	18:15	LB134493
	Lead	12.0	+/-12.0	U	12.0	P	01/30/2025	18:15	LB134493
	Selenium	20.0	+/-20.0	U	20.0	P	01/30/2025	18:15	LB134493
	Silver	10.0	+/-10.0	U	10.0	P	01/30/2025	18:15	LB134493
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	01/30/2025	18:51	LB134493
	Barium	100	+/-100	U	100	P	01/30/2025	18:51	LB134493
	Cadmium	6.00	+/-6.00	U	6.00	P	01/30/2025	18:51	LB134493
	Chromium	10.0	+/-10.0	U	10.0	P	01/30/2025	18:51	LB134493
	Lead	12.0	+/-12.0	U	12.0	P	01/30/2025	18:51	LB134493
	Selenium	20.0	+/-20.0	U	20.0	P	01/30/2025	18:51	LB134493
	Silver	10.0	+/-10.0	U	10.0	P	01/30/2025	18:51	LB134493