

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: RU2 Engineering, LLC **SDG No.:** Q1207
Contract: RUTW01 **Lab Code:** CHEM **Case No.:** Q1207 **SAS No.:** Q1207

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166342BS							
Arsenic	ug/L	4000	4010		100	80 - 120	P
Barium	ug/L	1000	889		89	80 - 120	P
Cadmium	ug/L	1000	955		96	80 - 120	P
Chromium	ug/L	2000	2050		102	80 - 120	P
Lead	ug/L	5000	4820		96	80 - 120	P
Selenium	ug/L	10000	9890		99	80 - 120	P
Silver	ug/L	380	372		98	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: RU2 Engineering, LLC

SDG No.: Q1207

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1207

SAS No.: Q1207

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166355BS Mercury	ug/L	4.0	4.51		113	80 - 120	CV