

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

Continuing Calib Date: 01/30/2025 Initial Calibration Date(s): 01/21/2025 01/22/2025

Continuing Calib Time: 06:10 Initial Calibration Time(s): 17:36 01:50

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.79	4.80	4.70	4.90	0.01
Aroclor-1016-2	(2)	4.81	4.81	4.71	4.91	0.00
Aroclor-1016-3	(3)	4.87	4.87	4.77	4.97	0.00
Aroclor-1016-4	(4)	4.99	4.99	4.89	5.09	0.00
Aroclor-1016-5	(5)	5.25	5.25	5.15	5.35	0.00
Aroclor-1260-1	(1)	6.29	6.29	6.19	6.39	0.00
Aroclor-1260-2	(2)	6.48	6.48	6.38	6.58	0.00
Aroclor-1260-3	(3)	6.85	6.85	6.75	6.95	0.00
Aroclor-1260-4	(4)	7.11	7.11	7.01	7.21	0.00
Aroclor-1260-5	(5)	7.35	7.35	7.25	7.45	0.00
Tetrachloro-m-xylene		3.70	3.70	3.60	3.80	0.00
Decachlorobiphenyl		8.76	8.76	8.66	8.86	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

Continuing Calib Date: 01/30/2025 Initial Calibration Date(s): 01/21/2025 01/22/2025

Continuing Calib Time: 06:10 Initial Calibration Time(s): 17:36 01:50

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.78	4.80	4.70	4.90	0.02
Aroclor-1016-2	(2)	4.80	4.82	4.72	4.92	0.02
Aroclor-1016-3	(3)	4.97	5.00	4.90	5.10	0.03
Aroclor-1016-4	(4)	5.02	5.04	4.94	5.14	0.02
Aroclor-1016-5	(5)	5.23	5.25	5.15	5.35	0.02
Aroclor-1260-1	(1)	6.26	6.29	6.19	6.39	0.03
Aroclor-1260-2	(2)	6.45	6.47	6.37	6.57	0.02
Aroclor-1260-3	(3)	6.60	6.63	6.53	6.73	0.03
Aroclor-1260-4	(4)	7.08	7.10	7.00	7.20	0.03
Aroclor-1260-5	(5)	7.32	7.34	7.24	7.44	0.02
Tetrachloro-m-xylene		3.70	3.72	3.62	3.82	0.02
Decachlorobiphenyl		8.71	8.73	8.63	8.83	0.02