

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL01 Date Analyzed: 01/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO109224.D Time Analyzed: 11:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.795	4.695	4.895	537.920	500.000	7.6
Aroclor-1016-2	4.815	4.714	4.914	537.190	500.000	7.4
Aroclor-1016-3	4.872	4.770	4.970	534.470	500.000	6.9
Aroclor-1016-4	4.993	4.891	5.091	539.610	500.000	7.9
Aroclor-1016-5	5.250	5.149	5.349	532.740	500.000	6.5
Aroclor-1260-1	6.294	6.190	6.390	523.850	500.000	4.8
Aroclor-1260-2	6.482	6.379	6.579	520.770	500.000	4.2
Aroclor-1260-3	6.851	6.747	6.947	524.120	500.000	4.8
Aroclor-1260-4	7.112	7.008	7.208	526.960	500.000	5.4
Aroclor-1260-5	7.353	7.249	7.449	528.070	500.000	5.6
Decachlorobiphenyl	8.764	8.658	8.858	52.620	50.000	5.2
Tetrachloro-m-xylene	3.700	3.600	3.800	53.800	50.000	7.6

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL01 Date Analyzed: 01/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO109224.D Time Analyzed: 11:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.782	4.703	4.903	562.120	500.000	12.4
Aroclor-1016-2	4.801	4.722	4.922	506.190	500.000	1.2
Aroclor-1016-3	4.976	4.898	5.098	526.530	500.000	5.3
Aroclor-1016-4	5.019	4.940	5.140	524.730	500.000	4.9
Aroclor-1016-5	5.232	5.153	5.353	522.620	500.000	4.5
Aroclor-1260-1	6.267	6.186	6.386	515.120	500.000	3.0
Aroclor-1260-2	6.454	6.373	6.573	508.150	500.000	1.6
Aroclor-1260-3	6.608	6.527	6.727	511.800	500.000	2.4
Aroclor-1260-4	7.080	6.999	7.199	511.760	500.000	2.4
Aroclor-1260-5	7.320	7.239	7.439	509.550	500.000	1.9
Decachlorobiphenyl	8.715	8.633	8.833	49.550	50.000	-0.9
Tetrachloro-m-xylene	3.697	3.618	3.818	54.550	50.000	9.1