

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL02 Date Analyzed: 01/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO109239.D Time Analyzed: 16:49

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.795	4.695	4.895	548.700	500.000	9.7
Aroclor-1016-2	4.815	4.714	4.914	549.830	500.000	10.0
Aroclor-1016-3	4.871	4.770	4.970	544.000	500.000	8.8
Aroclor-1016-4	4.992	4.891	5.091	546.500	500.000	9.3
Aroclor-1016-5	5.250	5.149	5.349	549.970	500.000	10.0
Aroclor-1260-1	6.292	6.190	6.390	541.430	500.000	8.3
Aroclor-1260-2	6.481	6.379	6.579	533.430	500.000	6.7
Aroclor-1260-3	6.850	6.747	6.947	540.540	500.000	8.1
Aroclor-1260-4	7.110	7.008	7.208	539.260	500.000	7.9
Aroclor-1260-5	7.352	7.249	7.449	539.520	500.000	7.9
Decachlorobiphenyl	8.763	8.658	8.858	53.320	50.000	6.6
Tetrachloro-m-xylene	3.700	3.600	3.800	54.500	50.000	9.0

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL02 Date Analyzed: 01/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO109239.D Time Analyzed: 16:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.781	4.703	4.903	523.890	500.000	4.8
Aroclor-1016-2	4.801	4.722	4.922	545.010	500.000	9.0
Aroclor-1016-3	4.976	4.898	5.098	534.620	500.000	6.9
Aroclor-1016-4	5.018	4.940	5.140	524.610	500.000	4.9
Aroclor-1016-5	5.231	5.153	5.353	535.200	500.000	7.0
Aroclor-1260-1	6.266	6.186	6.386	525.330	500.000	5.1
Aroclor-1260-2	6.453	6.373	6.573	521.480	500.000	4.3
Aroclor-1260-3	6.607	6.527	6.727	522.180	500.000	4.4
Aroclor-1260-4	7.079	6.999	7.199	522.690	500.000	4.5
Aroclor-1260-5	7.319	7.239	7.439	522.200	500.000	4.4
Decachlorobiphenyl	8.714	8.633	8.833	49.630	50.000	-0.7
Tetrachloro-m-xylene	3.697	3.618	3.818	54.990	50.000	10.0