

Surrogate Summary

SDG No.: Q1207

Client: RU2 Engineering, LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO108981.D	PIBLK-PO108981.D	Tetrachloro-m-xylene	1	20	21.9	109		60	140
		Decachlorobiphenyl	1	20	21.6	108		60	140
		Tetrachloro-m-xylene	2	20	20.6	103		60	140
		Decachlorobiphenyl	2	20	21.9	109		60	140
I.BLK-PO109228.D	PIBLK-PO109228.D	Tetrachloro-m-xylene	1	20	25.2	126		60	140
		Decachlorobiphenyl	1	20	25.4	127		60	140
		Tetrachloro-m-xylene	2	20	24.3	121		60	140
		Decachlorobiphenyl	2	20	24.0	120		60	140
PB166333BL	PB166333BL	Tetrachloro-m-xylene	1	20	25.2	126		32	144
		Decachlorobiphenyl	1	20	25.1	125		32	175
		Tetrachloro-m-xylene	2	20	24.1	120		32	144
		Decachlorobiphenyl	2	20	24.9	124		32	175
PB166333BS	PB166333BS	Tetrachloro-m-xylene	1	20	23.9	120		32	144
		Decachlorobiphenyl	1	20	25.7	129		32	175
		Tetrachloro-m-xylene	2	20	23.1	116		32	144
		Decachlorobiphenyl	2	20	24.0	120		32	175
I.BLK-PO109243.D	PIBLK-PO109243.D	Tetrachloro-m-xylene	1	20	24.6	123		60	140
		Decachlorobiphenyl	1	20	25.3	126		60	140
		Tetrachloro-m-xylene	2	20	23.7	118		60	140
		Decachlorobiphenyl	2	20	23.6	118		60	140
Q1205-01MS	VNJ-236MS	Tetrachloro-m-xylene	1	20	19.5	98		32	144
		Decachlorobiphenyl	1	20	16.1	80		32	175
		Tetrachloro-m-xylene	2	20	18.9	94		32	144
		Decachlorobiphenyl	2	20	15.7	79		32	175
Q1205-01MSD	VNJ-236MSD	Tetrachloro-m-xylene	1	20	18.8	94		32	144
		Decachlorobiphenyl	1	20	15.4	77		32	175
		Tetrachloro-m-xylene	2	20	18.5	92		32	144
		Decachlorobiphenyl	2	20	15.5	77		32	175
Q1207-03	JPP-2.1-012725	Tetrachloro-m-xylene	1	20	22.1	110		32	144
		Decachlorobiphenyl	1	20	13.2	66		32	175
		Tetrachloro-m-xylene	2	20	19.8	99		32	144
		Decachlorobiphenyl	2	20	13.9	69		32	175
Q1207-07	JPP-5.1-012725	Tetrachloro-m-xylene	1	20	18.9	95		32	144
		Decachlorobiphenyl	1	20	14.1	70		32	175
		Tetrachloro-m-xylene	2	20	20.6	103		32	144
		Decachlorobiphenyl	2	20	14.8	74		32	175
Q1207-11	JPP-4.5-012725	Tetrachloro-m-xylene	1	20	16.1	80		32	144
		Decachlorobiphenyl	1	20	10.7	53		32	175
		Tetrachloro-m-xylene	2	20	16.6	83		32	144
		Decachlorobiphenyl	2	20	11.4	57		32	175
Q1207-15	JPP-16.2-012725	Tetrachloro-m-xylene	1	20	19.0	95		32	144

Surrogate Summary

SDG No.: Q1207

Client: RU2 Engineering, LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1207-15	JPP-16.2-012725	Decachlorobiphenyl	1	20	13.5	68		32	175
		Tetrachloro-m-xylene	2	20	20.1	100		32	144
		Decachlorobiphenyl	2	20	14.3	72		32	175
Q1207-19	JPP-20.2-012725	Tetrachloro-m-xylene	1	20	19.6	98		32	144
		Decachlorobiphenyl	1	20	11.1	56		32	175
		Tetrachloro-m-xylene	2	20	20.0	100		32	144
		Decachlorobiphenyl	2	20	12.0	60		32	175
I.BLK-PO109258.D	PIBLK-PO109258.D	Tetrachloro-m-xylene	1	20	23.4	117		60	140
		Decachlorobiphenyl	1	20	18.9	94		60	140
		Tetrachloro-m-xylene	2	20	23.9	119		60	140
		Decachlorobiphenyl	2	20	20.6	103		60	140