

Surrogate Summary

SDG No.: Q1207

Client: RU2 Engineering, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL093725.D	PIBLK-PL093725.D	Decachlorobiphenyl	1	20	22.1	111		43	140
		Tetrachloro-m-xylene	1	20	20.8	104		77	126
		Decachlorobiphenyl	2	20	21.9	109		43	140
		Tetrachloro-m-xylene	2	20	20.5	103		77	126
I.BLK-PL093874.D	PIBLK-PL093874.D	Decachlorobiphenyl	1	20	22.5	113		43	140
		Tetrachloro-m-xylene	1	20	19.9	100		77	126
		Decachlorobiphenyl	2	20	21.0	105		43	140
		Tetrachloro-m-xylene	2	20	18.1	91		77	126
Q1209-05MS	WC-5MS	Decachlorobiphenyl	1	20	22.2	111		10	148
		Tetrachloro-m-xylene	1	20	20.2	101		10	159
		Decachlorobiphenyl	2	20	21.5	107		10	148
		Tetrachloro-m-xylene	2	20	19.4	97		10	159
Q1209-05MSD	WC-5MSD	Decachlorobiphenyl	1	20	22.2	111		10	148
		Tetrachloro-m-xylene	1	20	20.4	102		10	159
		Decachlorobiphenyl	2	20	21.6	108		10	148
		Tetrachloro-m-xylene	2	20	19.8	99		10	159
PB166334BL	PB166334BL	Decachlorobiphenyl	1	20	24.8	124		10	148
		Tetrachloro-m-xylene	1	20	22.0	110		10	159
		Decachlorobiphenyl	2	20	24.4	122		10	148
		Tetrachloro-m-xylene	2	20	20.6	103		10	159
PB166334BS	PB166334BS	Decachlorobiphenyl	1	20	19.6	98		10	148
		Tetrachloro-m-xylene	1	20	16.9	84		10	159
		Decachlorobiphenyl	2	20	19.0	95		10	148
		Tetrachloro-m-xylene	2	20	15.9	80		10	159
I.BLK-PL093887.D	PIBLK-PL093887.D	Decachlorobiphenyl	1	20	21.4	107		43	140
		Tetrachloro-m-xylene	1	20	18.4	92		77	126
		Decachlorobiphenyl	2	20	21.2	106		43	140
		Tetrachloro-m-xylene	2	20	17.5	88		77	126
Q1207-03	JPP-2.1-012725	Decachlorobiphenyl	1	20	11.2	56		10	148
		Tetrachloro-m-xylene	1	20	12.5	63		10	159
		Decachlorobiphenyl	2	20	9.33	47		10	148
		Tetrachloro-m-xylene	2	20	11.3	56		10	159
Q1207-07	JPP-5.1-012725	Decachlorobiphenyl	1	20	15.0	75		10	148
		Tetrachloro-m-xylene	1	20	13.3	67		10	159
		Decachlorobiphenyl	2	20	10.5	53		10	148
		Tetrachloro-m-xylene	2	20	14.3	72		10	159
Q1207-11	JPP-4.5-012725	Decachlorobiphenyl	1	20	10.5	53		10	148
		Tetrachloro-m-xylene	1	20	11.8	59		10	159
		Decachlorobiphenyl	2	20	7.97	40		10	148
		Tetrachloro-m-xylene	2	20	12.2	61		10	159
Q1207-15	JPP-16.2-012725	Decachlorobiphenyl	1	20	13.9	70		10	148

Surrogate Summary

SDG No.: Q1207

Client: RU2 Engineering, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1207-15	JPP-16.2-012725	Tetrachloro-m-xylene	1	20	13.5	67		10	159
		Decachlorobiphenyl	2	20	9.81	49		10	148
		Tetrachloro-m-xylene	2	20	13.7	68		10	159
Q1207-19	JPP-20.2-012725	Decachlorobiphenyl	1	20	11.1	55		10	148
		Tetrachloro-m-xylene	1	20	12.6	63		10	159
		Decachlorobiphenyl	2	20	8.50	43		10	148
		Tetrachloro-m-xylene	2	20	12.1	60		10	159
		Decachlorobiphenyl	1	20	20.2	101		43	140
I.BLK-PL093904.D	PIBLK-PL093904.D	Tetrachloro-m-xylene	1	20	19.1	96		77	126
		Decachlorobiphenyl	2	20	16.0	80		43	140
		Tetrachloro-m-xylene	2	20	18.6	93		77	126
		Decachlorobiphenyl	1	20	16.0	80		43	140