

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM **Case No.:** Q1207 **SAS No.:** Q1207 **SDG NO.:** Q1207

Continuing Calib Date: 01/31/2025 **Initial Calibration Date(s):** 01/14/2025 01/14/2025

Continuing Calib Time: 01:01 **Initial Calibration Time(s):** 10:31 12:07

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.20	7.20	7.10	7.30	0.00
2,4-D	8.31	8.32	8.22	8.42	0.01
2,4,5-TP(Silvex)	9.19	9.19	9.09	9.29	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM **Case No.:** Q1207 **SAS No.:** Q1207 **SDG NO.:** Q1207

Continuing Calib Date: 01/31/2025 **Initial Calibration Date(s):** 01/14/2025 01/14/2025

Continuing Calib Time: 01:01 **Initial Calibration Time(s):** 10:31 12:07

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.67	7.68	7.58	7.78	0.02
2,4-D	8.90	8.91	8.81	9.01	0.01
2,4,5-TP(Silvex)	9.80	9.81	9.71	9.91	0.01