

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL01 Date Analyzed: 01/30/2025

Lab Sample No.: HSTDCCC750 Data File : PS028989.D Time Analyzed: 13:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.181	9.089	9.289	750.680	712.500	5.4
2,4-D	8.310	8.216	8.416	735.240	705.000	4.3
2,4-DCAA	7.192	7.097	7.297	777.790	750.000	3.7

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL01 Date Analyzed: 01/30/2025

Lab Sample No.: HSTDCCC750 Data File : PS028989.D Time Analyzed: 13:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.803	9.713	9.913	766.570	712.500	7.6
2,4-D	8.905	8.813	9.013	711.080	705.000	0.9
2,4-DCAA	7.670	7.577	7.777	736.960	750.000	-1.7