

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL02 Date Analyzed: 01/30/2025

Lab Sample No.: HSTDCCC750 Data File : PS028998.D Time Analyzed: 16:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.181	9.089	9.289	751.030	712.500	5.4
2,4-D	8.309	8.216	8.416	735.060	705.000	4.3
2,4-DCAA	7.192	7.097	7.297	774.330	750.000	3.2

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL02 Date Analyzed: 01/30/2025

Lab Sample No.: HSTDCCC750 Data File : PS028998.D Time Analyzed: 16:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.803	9.713	9.913	780.240	712.500	9.5
2,4-D	8.904	8.813	9.013	724.010	705.000	2.7
2,4-DCAA	7.670	7.577	7.777	751.950	750.000	0.3