

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL03 Date Analyzed: 01/31/2025

Lab Sample No.: HSTDCCC750 Data File : PS029005.D Time Analyzed: 01:01

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.185	9.089	9.289	773.450	712.500	8.6
2,4-D	8.313	8.216	8.416	756.410	705.000	7.3
2,4-DCAA	7.195	7.097	7.297	802.310	750.000	7.0

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL03 Date Analyzed: 01/31/2025

Lab Sample No.: HSTDCCC750 Data File : PS029005.D Time Analyzed: 01:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.798	9.713	9.913	805.070	712.500	13.0
2,4-D	8.899	8.813	9.013	745.580	705.000	5.8
2,4-DCAA	7.665	7.577	7.777	790.770	750.000	5.4