

Surrogate Summary

SDG No.: Q1207

Client: RU2 Engineering, LLC

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PS028900.D	PIBLK-PS028900.D	2,4-DCAA	1	500	474	95		39	175
		2,4-DCAA	2	500	492	98		39	175
I.BLK-PS028988.D	PIBLK-PS028988.D	2,4-DCAA	1	500	512	102		39	175
		2,4-DCAA	2	500	489	98		39	175
PB166382BL	PB166382BL	2,4-DCAA	1	500	445	89		39	175
		2,4-DCAA	2	500	404	81		39	175
PB166382BS	PB166382BS	2,4-DCAA	1	500	510	102		39	175
		2,4-DCAA	2	500	465	93		39	175
PB166318TB	PB166318TB	2,4-DCAA	1	500	462	92		39	175
		2,4-DCAA	2	500	313	63		39	175
Q1206-04MS	JPP-20.1-012725MS	2,4-DCAA	1	500	424	85		39	175
		2,4-DCAA	2	500	267	53		39	175
Q1206-04MSD	JPP-20.1-012725MSD	2,4-DCAA	1	500	424	85		39	175
		2,4-DCAA	2	500	265	53		39	175
I.BLK-PS028997.D	PIBLK-PS028997.D	2,4-DCAA	1	500	510	102		39	175
		2,4-DCAA	2	500	499	100		39	175
Q1207-04	JPP-2.1-012725	2,4-DCAA	1	500	373	75		39	175
		2,4-DCAA	2	500	275	55		39	175
Q1207-08	JPP-5.1-012725	2,4-DCAA	1	500	380	76		39	175
		2,4-DCAA	2	500	258	52		39	175
Q1207-12	JPP-4.5-012725	2,4-DCAA	1	500	460	92		39	175
		2,4-DCAA	2	500	326	65		39	175
Q1207-16	JPP-16.2-012725	2,4-DCAA	1	500	514	103		39	175
		2,4-DCAA	2	500	352	70		39	175
Q1207-20	JPP-20.2-012725	2,4-DCAA	1	500	497	99		39	175
		2,4-DCAA	2	500	346	69		39	175
I.BLK-PS029004.D	PIBLK-PS029004.D	2,4-DCAA	1	500	511	102		39	175
		2,4-DCAA	2	500	502	100		39	175