

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

Continuing Calib Date: 01/30/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 13:53 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.06	9.05	8.95	9.15	-0.01
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.92	4.91	4.81	5.01	-0.01
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

Continuing Calib Date: 01/30/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 13:53 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.91	7.91	7.81	8.01	0.00
Tetrachloro-m-xylene	2.77	2.77	2.67	2.87	0.00
gamma-BHC (Lindane)	3.61	3.61	3.51	3.71	0.00
Heptachlor	3.95	3.95	3.85	4.05	0.01
Heptachlor epoxide	4.73	4.73	4.63	4.83	0.00
Endrin	5.64	5.64	5.54	5.74	0.00
Methoxychlor	6.61	6.61	6.51	6.71	0.00