

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 01/30/2025

Lab Sample No.: PSTDCCC050 Data File : PL093876.D Time Analyzed: 09:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.055	8.953	9.153	48.590	50.000	-2.8
Endrin	6.574	6.472	6.672	48.110	50.000	-3.8
gamma-BHC (Lindane)	4.327	4.227	4.427	45.140	50.000	-9.7
Heptachlor	4.915	4.814	5.014	47.560	50.000	-4.9
Heptachlor epoxide	5.683	5.582	5.782	45.130	50.000	-9.7
Methoxychlor	7.500	7.398	7.598	51.550	50.000	3.1
Tetrachloro-m-xylene	3.539	3.439	3.639	45.390	50.000	-9.2

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 01/30/2025

Lab Sample No.: PSTDCCC050 Data File : PL093876.D Time Analyzed: 09:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.910	7.810	8.010	49.800	50.000	-0.4
Endrin	5.636	5.536	5.736	48.870	50.000	-2.3
gamma-BHC (Lindane)	3.607	3.507	3.707	46.010	50.000	-8.0
Heptachlor	3.945	3.845	4.045	47.440	50.000	-5.1
Heptachlor epoxide	4.727	4.627	4.827	46.460	50.000	-7.1
Methoxychlor	6.610	6.509	6.709	49.830	50.000	-0.3
Tetrachloro-m-xylene	2.774	2.674	2.874	45.470	50.000	-9.1