

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL04 Date Analyzed: 01/30/2025

Lab Sample No.: PSTDCCC050 Data File : PL093888.D Time Analyzed: 13:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.056	8.953	9.153	48.490	50.000	-3.0
Endrin	6.573	6.472	6.672	43.820	50.000	-12.4
gamma-BHC (Lindane)	4.328	4.227	4.427	44.710	50.000	-10.6
Heptachlor	4.916	4.814	5.014	46.250	50.000	-7.5
Heptachlor epoxide	5.684	5.582	5.782	44.430	50.000	-11.1
Methoxychlor	7.500	7.398	7.598	46.980	50.000	-6.0
Tetrachloro-m-xylene	3.539	3.439	3.639	45.400	50.000	-9.2

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL04 Date Analyzed: 01/30/2025

Lab Sample No.: PSTDCCC050 Data File : PL093888.D Time Analyzed: 13:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.910	7.810	8.010	48.940	50.000	-2.1
Endrin	5.637	5.536	5.736	45.280	50.000	-9.4
gamma-BHC (Lindane)	3.607	3.507	3.707	44.600	50.000	-10.8
Heptachlor	3.945	3.845	4.045	45.400	50.000	-9.2
Heptachlor epoxide	4.727	4.627	4.827	44.970	50.000	-10.1
Methoxychlor	6.610	6.509	6.709	45.430	50.000	-9.1
Tetrachloro-m-xylene	2.774	2.674	2.874	44.780	50.000	-10.4