

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL05 Date Analyzed: 01/30/2025

Lab Sample No.: PSTDCCC050 Data File : PL093906.D Time Analyzed: 19:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.053	8.953	9.153	45.970	50.000	-8.1
Endrin	6.571	6.472	6.672	41.650	50.000	-16.7
gamma-BHC (Lindane)	4.326	4.227	4.427	45.020	50.000	-10.0
Heptachlor	4.914	4.814	5.014	45.190	50.000	-9.6
Heptachlor epoxide	5.681	5.582	5.782	43.240	50.000	-13.5
Methoxychlor	7.498	7.398	7.598	44.720	50.000	-10.6
Tetrachloro-m-xylene	3.537	3.439	3.639	45.460	50.000	-9.1

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL05 Date Analyzed: 01/30/2025

Lab Sample No.: PSTDCCC050 Data File : PL093906.D Time Analyzed: 19:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.909	7.810	8.010	45.120	50.000	-9.8
Endrin	5.635	5.536	5.736	44.130	50.000	-11.7
gamma-BHC (Lindane)	3.606	3.507	3.707	46.060	50.000	-7.9
Heptachlor	3.944	3.845	4.045	45.820	50.000	-8.4
Heptachlor epoxide	4.726	4.627	4.827	46.510	50.000	-7.0
Methoxychlor	6.609	6.509	6.709	42.130	50.000	-15.7
Tetrachloro-m-xylene	2.774	2.674	2.874	46.540	50.000	-6.9