

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

EPA Sample No.: SSTDCCC040 Date Analyzed: 02/03/2025

Lab File ID: BF141373.D Time Analyzed: 17:14

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	104167	6.798	414412	8.08	226326	9.83
UPPER LIMIT	208334	7.298	828824	8.581	452652	10.333
LOWER LIMIT	52083.5	6.298	207206	7.581	113163	9.333
EPA SAMPLE NO.						
01 PB166335BL	90939	6.79	365110	8.08	206122	9.83

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.