

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

EPA Sample No.: SSTDCCC040 Date Analyzed: 02/05/2025

Lab File ID: BF141447.D Time Analyzed: 15:16

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	293274	11.316	189132	13.963	170936	15.421
UPPER LIMIT	586548	11.816	378264	14.463	341872	15.921
LOWER LIMIT	146637	10.816	94566	13.463	85468	14.921
EPA SAMPLE NO.						
01 JPP-2.1-012725	175054	11.31	157036	13.96	119629	15.41
02 JPP-5.1-012725	162078	11.32	126468	13.96	93403	15.42
03 JPP-4.5-012725	168524	11.32	142671	13.96	103713	15.42
04 JPP-16.2-012725	186178	11.32	135658	13.96	100723	15.42
05 JPP-20.2-012725	194970	11.31	168850	13.96	153896	15.42

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.