

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: RUTW01

Lab Code: CHEM

SAS No.: Q1207 SDG NO.: Q1207

Lab File ID: BF141300.D

DFTPP Injection Date: 01/30/2025

Instrument ID: BNA_F

DFTPP Injection Time: 08:28

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	45
68	Less than 2.0% of mass 69	0.7 (1.9) 1
69	Mass 69 relative abundance	37.4
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	49.1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	25.8
365	Greater than 1% of mass 198	3
441	Present, but less than mass 443	11.5
442	Greater than 50% of mass 198	72.8
443	15.0 - 24.0% of mass 442	14.4 (19.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF141301.D	01/30/2025	08:54
PB166335BS	PB166335BS	BF141303.D	01/30/2025	09:57
WC-5MS	Q1209-05MS	BF141315.D	01/30/2025	16:03
WC-5MSD	Q1209-05MSD	BF141316.D	01/30/2025	16:30