

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1207

Client: RU2 Engineering, LLC

Analytical Method: 8270E

DataFile: BF141445.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB166354BS	Pyridine	50	38.6	ug/L	77				29	97	
	1,4-Dichlorobenzene	50	46.7	ug/L	93				76	103	
	2-Methylphenol	50	45.8	ug/L	92				69	109	
	3+4-Methylphenols	50	48.2	ug/L	96				67	106	
	Hexachloroethane	50	48.9	ug/L	98				76	118	
	Nitrobenzene	50	47.0	ug/L	94				58	106	
	Hexachlorobutadiene	50	49.7	ug/L	99				69	101	
	2,4,6-Trichlorophenol	50	50.3	ug/L	101				61	110	
	2,4,5-Trichlorophenol	50	49.8	ug/L	100				70	106	
	2,4-Dinitrotoluene	50	52.3	ug/L	105				60	115	
	Hexachlorobenzene	50	49.8	ug/L	100				73	106	
	Pentachlorophenol	100	100	ug/L	100				47	114	