

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207

EPA Sample No.: SSTDCCC040 Date Analyzed: 01/31/2025

Lab File ID: BF141336.D Time Analyzed: 17:19

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	140033	6.798	558386	8.08	299701	9.84
UPPER LIMIT	280066	7.298	1116770	8.581	599402	10.339
LOWER LIMIT	70016.5	6.298	279193	7.581	149851	9.339
EPA SAMPLE NO.						
01 PB166354BL	138010	6.80	554699	8.08	302564	9.83
02 VNJ-236MS	142849	6.80	564006	8.08	294756	9.84
03 VNJ-236MSD	153925	6.80	605647	8.08	316012	9.84

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.