

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG No.: Q1207

Instrument ID: MSVOA_N Calibration Date/Time: 01/29/2025 11:08

Lab File ID: VN085548.D Init. Calib. Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 14:56 17:19

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.737	0.641		-13.03	20
1,1-Dichloroethene	0.537	0.511		-4.84	20
2-Butanone	0.384	0.403		4.95	20
Carbon Tetrachloride	0.557	0.568		1.98	20
Chloroform	1.218	1.236		1.48	20
Benzene	1.463	1.501		2.6	20
1,2-Dichloroethane	0.551	0.584		5.99	20
Trichloroethene	0.341	0.338		-0.88	20
Tetrachloroethene	0.341	0.339		-0.59	20
Chlorobenzene	1.095	1.109	0.3	1.28	20
1,2-Dichloroethane-d4	0.807	0.877		8.67	20
Dibromofluoromethane	0.347	0.387		11.53	20
Toluene-d8	1.232	1.409		14.37	20
4-Bromofluorobenzene	0.422	0.503		19.19	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.