

Surrogate Summary

SDG No.: Q1207

Client: RU2 Engineering, LLC

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1207-02	JPP-2.1-012725	1,2-Dichloroethane-d4	50	57.3	115	74	125
		Dibromofluoromethane	50	51.0	102	75	124
		Toluene-d8	50	52.5	105	86	113
		4-Bromofluorobenzene	50	44.6	89	77	121
Q1207-06	JPP-5.1-012725	1,2-Dichloroethane-d4	50	57.0	114	74	125
		Dibromofluoromethane	50	52.2	104	75	124
		Toluene-d8	50	52.8	106	86	113
		4-Bromofluorobenzene	50	44.3	89	77	121
Q1207-10	JPP-4.5-012725	1,2-Dichloroethane-d4	50	57.3	115	74	125
		Dibromofluoromethane	50	52.9	106	75	124
		Toluene-d8	50	53.1	106	86	113
		4-Bromofluorobenzene	50	46.6	93	77	121
Q1207-14	JPP-16.2-012725	1,2-Dichloroethane-d4	50	57.5	115	74	125
		Dibromofluoromethane	50	52.5	105	75	124
		Toluene-d8	50	53.3	107	86	113
		4-Bromofluorobenzene	50	44.0	88	77	121
Q1207-18	JPP-20.2-012725	1,2-Dichloroethane-d4	50	58.3	117	74	125
		Dibromofluoromethane	50	51.7	103	75	124
		Toluene-d8	50	53.5	107	86	113
		4-Bromofluorobenzene	50	44.3	89	77	121
VN0129WBL01	VN0129WBL01	1,2-Dichloroethane-d4	50	57.3	115	74	125
		Dibromofluoromethane	50	53.7	107	75	124
		Toluene-d8	50	50.2	100	86	113
		4-Bromofluorobenzene	50	48.3	97	77	121
VN0129WBS01	VN0129WBS01	1,2-Dichloroethane-d4	50	50.8	102	74	125
		Dibromofluoromethane	50	52.1	104	75	124
		Toluene-d8	50	53.0	106	86	113
		4-Bromofluorobenzene	50	55.3	111	77	121
VN0129WBSD0	VN0129WBSD01	1,2-Dichloroethane-d4	50	52.0	104	74	125
		Dibromofluoromethane	50	52.8	106	75	124
		Toluene-d8	50	53.1	106	86	113
		4-Bromofluorobenzene	50	54.8	110	77	121