

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1207

Client: RU2 Engineering, LLC

Analytical Method: SW8260D

Datafile : VN085551.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0129WBS01	Vinyl chloride	20	16.6	ug/L	83			65	117	
	1,1-Dichloroethene	20	18.0	ug/L	90			74	110	
	2-Butanone	100	100	ug/L	100			65	122	
	Carbon Tetrachloride	20	19.5	ug/L	98			77	113	
	Chloroform	20	19.2	ug/L	96			79	113	
	Benzene	20	19.4	ug/L	97			82	109	
	1,2-Dichloroethane	20	20.0	ug/L	100			80	115	
	Trichloroethene	20	18.9	ug/L	95			77	113	
	Tetrachloroethene	20	19.4	ug/L	97			67	123	
	Chlorobenzene	20	19.3	ug/L	97			82	109	