

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** Q1207

**Client:** RU2 Engineering, LLC

**Analytical Method:** SW8260D

**Datafile :** VN085552.D

| Lab Sample ID | Parameter            | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits<br>High | RPD |
|---------------|----------------------|-------|--------|------|-----|-----|------|-----|----------------|-----|
| VN0129WBSD01  | Vinyl chloride       | 20    | 16.3   | ug/L | 81  | 2   |      | 65  | 117            | 20  |
|               | 1,1-Dichloroethene   | 20    | 17.9   | ug/L | 90  | 0   |      | 74  | 110            | 20  |
|               | 2-Butanone           | 100   | 110    | ug/L | 110 | 10  |      | 65  | 122            | 20  |
|               | Carbon Tetrachloride | 20    | 18.8   | ug/L | 94  | 4   |      | 77  | 113            | 20  |
|               | Chloroform           | 20    | 19.4   | ug/L | 97  | 1   |      | 79  | 113            | 20  |
|               | Benzene              | 20    | 19.3   | ug/L | 97  | 0   |      | 82  | 109            | 20  |
|               | 1,2-Dichloroethane   | 20    | 20.6   | ug/L | 103 | 3   |      | 80  | 115            | 20  |
|               | Trichloroethene      | 20    | 18.3   | ug/L | 92  | 3   |      | 77  | 113            | 20  |
|               | Tetrachloroethene    | 20    | 19.1   | ug/L | 96  | 1   |      | 67  | 123            | 20  |
|               | Chlorobenzene        | 20    | 19.2   | ug/L | 96  | 1   |      | 82  | 109            | 20  |