

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG No.: Q1207

Instrument ID: MSVOA_Y Calibration Date/Time: 02/04/2025 09:47

Lab File ID: VY021046.D Init. Calib. Date(s): 02/03/2025 02/03/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:35 12:44

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.446	0.470		5.38	20
Chloromethane	0.440	0.443	0.1	0.68	20
Vinyl Chloride	0.443	0.443		0	20
Bromomethane	0.274	0.270		-1.46	20
Chloroethane	0.263	0.263		0	20
Trichlorofluoromethane	0.806	0.812		0.74	20
1,1,2-Trichlorotrifluoroethane	0.506	0.522		3.16	20
Tert butyl alcohol	0.035	0.035		0	20
1,1-Dichloroethene	0.477	0.489		2.52	20
Acetone	0.080	0.085		6.25	20
Carbon Disulfide	1.495	1.571		5.08	20
Methyl tert-butyl Ether	1.205	1.270		5.39	20
Methyl Acetate	0.280	0.312		11.43	20
Methylene Chloride	0.497	0.513		3.22	20
trans-1,2-Dichloroethene	0.525	0.550		4.76	20
1,1-Dichloroethane	0.970	1.014	0.1	4.54	20
Cyclohexane	0.874	0.883		1.03	20
2-Butanone	0.131	0.143		9.16	20
Carbon Tetrachloride	0.555	0.591		6.49	20
cis-1,2-Dichloroethene	0.604	0.639		5.8	20
Bromochloromethane	0.384	0.407		5.99	20
Chloroform	0.997	1.045		4.81	20
1,1,1-Trichloroethane	0.927	0.969		4.53	20
Methylcyclohexane	0.581	0.623		7.23	20
Benzene	1.393	1.501		7.75	20
1,2-Dichloroethane	0.386	0.414		7.25	20
Trichloroethene	0.360	0.380		5.56	20
1,2-Dichloropropane	0.331	0.357		7.86	20
Bromodichloromethane	0.491	0.518		5.5	20
4-Methyl-2-Pentanone	0.208	0.238		14.42	20
Toluene	0.875	0.954		9.03	20
t-1,3-Dichloropropene	0.441	0.486		10.2	20
cis-1,3-Dichloropropene	0.522	0.569		9	20
1,1,2-Trichloroethane	0.237	0.253		6.75	20
2-Hexanone	0.134	0.157		17.16	20
Dibromochloromethane	0.333	0.359		7.81	20
1,2-Dibromoethane	0.222	0.240		8.11	20
Tetrachloroethene	0.371	0.387		4.31	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG No.: Q1207

Instrument ID: MSVOA_Y Calibration Date/Time: 02/04/2025 09:47

Lab File ID: VY021046.D Init. Calib. Date(s): 02/03/2025 02/03/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:35 12:44

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chlorobenzene	1.113	1.159	0.3	4.13	20
Ethyl Benzene	1.969	2.119		7.62	20
m/p-Xylenes	0.737	0.793		7.6	20
o-Xylene	0.687	0.747		8.73	20
Styrene	1.145	1.261		10.13	20
Bromoform	0.220	0.240	0.1	9.09	20
Isopropylbenzene	3.906	4.054		3.76	20
1,1,2,2-Tetrachloroethane	0.650	0.677	0.3	4.15	20
1,3-Dichlorobenzene	1.740	1.821		4.66	20
1,4-Dichlorobenzene	1.714	1.787		4.26	20
1,2-Dichlorobenzene	1.510	1.568		3.84	20
1,2-Dibromo-3-Chloropropane	0.097	0.101		4.12	20
1,2,4-Trichlorobenzene	0.841	0.932		10.82	20
1,2,3-Trichlorobenzene	0.700	0.777		11	20
1,2-Dichloroethane-d4	0.512	0.534		4.3	20
Dibromofluoromethane	0.320	0.343		7.19	20
Toluene-d8	1.220	1.329		8.93	20
4-Bromofluorobenzene	0.399	0.446		11.78	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.