

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RUTW01
 Lab Code: CHEM Case No.: Q1207 SAS No.: Q1207 SDG NO.: Q1207
 Lab File ID: VY021023.D Date Analyzed: 02/03/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:43
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	121069	13.353				
UPPER LIMIT	242138	13.853				
LOWER LIMIT	60534.5	12.853				
EPA SAMPLE NO.						
JPP-2.1-012725	69823	13.35				
JPP-5.1-012725	120538	13.35				
JPP-4.5-012725	103960	13.35				
VY0203SBL01	89793	13.35				
VY0203SBS01	127785	13.35				
VY0203SBSD01	128240	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.