

ANALYTICAL RESULTS SUMMARY

VOLATILE ORGANICS
GENERAL CHEMISTRY
METALS
GC SEMI-VOLATILES
SEMI-VOLATILE ORGANICS

PROJECT NAME : FT MEADE TIPTON AIRFIELD PARCEL RI - PO 0111169

WESTON SOLUTIONS

1400 Weston Way

PO Box 2653

West Chester, PA - 19380

Phone No: 610-701-7400

ORDER ID : Q1211

ATTENTION : Nathan Fretz



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	4
2.1) VOC-TCLVOA-10- Case Narrative	4
2.2) Gasoline Range Organics- Case Narrative	6
2.3) SVOC-TCL BNA -20- Case Narrative	8
2.4) PESTICIDE Group1- Case Narrative	10
2.5) Diesel Range Organics- Case Narrative	12
2.6) Metals-MS- Case Narrative	14
2.7) Genchem- Case Narrative	16
3) Qualifier Page	18
4) QA Checklist	20
5) VOC-TCLVOA-10 Data	21
6) Gasoline Range Organics Data	36
7) SVOC-TCL BNA -20 Data	40
8) PESTICIDE Group1 Data	51
9) Diesel Range Organics Data	58
10) Metals-MS Data	62
11) Genchem Data	67
12) Shipping Document	72
12.1) CHAIN OF CUSTODY	73
12.2) Lab Certificate	74
12.3) Internal COC	75

1
2
3
4
5
6
7
8
9
10
11
12

Cover Page

Order ID : Q1211

Project ID : Ft Meade Tipton Airfield Parcel RI - PO 0111169

Client : Weston Solutions

Lab Sample Number

Q1211-01
Q1211-02
Q1211-03

Client Sample Number

TAPHHA-MW01-012825-00-T4
TAPIAL2-MW03-012825-00-T3
TAP-TB-02-012825-T4

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 2/11/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Weston Solutions

Project Name: Ft Meade Tipton Airfield Parcel RI - PO 0111169

Project # N/A

Chemtech Project # Q1211

Test Name: VOC-TCLVOA-10

A. Number of Samples and Date of Receipt:

3 Water samples were received on 01/29/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Ammonia, Anions Group5, Diesel Range Organics, Gasoline Range Organics, Hardness, Total, Hexavalent Chromium, Mercury, Metals ICP-TAL, METALS-TAL, Oil and Grease, PESTICIDE Group1, SVOC-TCL BNA -20, TOC and VOC-TCLVOA-10. This data package contains results for VOC-TCLVOA-10.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_N were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868. The analysis of VOC-TCLVOA-10 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for TAPIAL2-MW03-012825-00-T3 [4-Bromofluorobenzene - 83%] as a corrective action sample reanalyzed and both run reported.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration File ID VN085576.D met the requirements except for Carbon Disulfide failing marginally low and 2-Hexanone,4-Methyl-2-Pentanone failing high but no positive hit in associated samples therefore no corrective action taken.

The Continuous Calibration File ID VN085591.D met the requirements except for Chloromethane failing marginally low and Carbon Disulfide failing low associated

sample initially analyzed in sequence VN013025 where surrogate failing as a corrective action sample reanalyzed but CCAL failing low however no positive hit in associated sample therefore both run reported.

The Tuning criteria met requirements.

E. Additional Comments:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Calculation for Concentration in Water Samples:

$$\text{Concentration ug/L} = \frac{(A_x)(I_s)(Df)}{(A_{is})(RRF)(V_0)}$$

Where,

A_x = Area for the compound to be measured

A_{is} = Area for the specific internal standard

I_s = Amount of internal standard added in nanograms (ng)

RRF = Relative response factor of the initial calibration curve standard.

V₀ = Volume of water purged in milliliters (mL)

Df = Dilution factor.

G. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature _____

CASE NARRATIVE

Weston Solutions

Project Name: Ft Meade Tipton Airfield Parcel RI - PO 0111169

Project # N/A

Chemtech Project # Q1211

Test Name: Gasoline Range Organics

A. Number of Samples and Date of Receipt:

3 Water samples were received on 01/29/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Ammonia, Anions Group5, Diesel Range Organics, Gasoline Range Organics, Hardness, Total, Hexavalent Chromium, Mercury, Metals ICP-TAL, METALS-TAL, Oil and Grease, PESTICIDE Group1, SVOC-TCL BNA -20, TOC and VOC-TCLVOA-10. This data package contains results for Gasoline Range Organics.

C. Analytical Techniques:

The analysis performed on instrument FID_B were done using GC column RTX502.2 which is 60 meters, 0.53mm ID, 3.0 um df, cat#10909. The analysis of Gasoline Range Organics was based on method 8015D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

E. Additional Comments:

The not QT review data is reported in the Miscellaneous.

F. Calculation for Concentration in WATER samples :

Calculations for samples are: Waters:

$$\text{mg/L} = \frac{\text{ng purged}}{(\text{mL sample purged}) (1000)}$$

Where

$$\text{ng purged} = \frac{\text{total area of peaks}}{\text{calibration factor (CF)}}$$

CF = mean CF of the initial calibration

G. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

CASE NARRATIVE

Weston Solutions

Project Name: Ft Meade Tipton Airfield Parcel RI - PO 0111169

Project # N/A

Chemtech Project # Q1211

Test Name: SVOC-TCL BNA -20

A. Number of Samples and Date of Receipt:

3 Water samples were received on 01/29/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Ammonia, Anions Group5, Diesel Range Organics, Gasoline Range Organics, Hardness, Total, Hexavalent Chromium, Mercury, Metals ICP-TAL, METALS-TAL, Oil and Grease, PESTICIDE Group1, SVOC-TCL BNA -20, TOC and VOC-TCLVOA-10. This data package contains results for SVOC-TCL BNA -20.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um dfThe analysis of SVOC-TCL BNA -20 was based on method 8270E and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike for {PB166363BS} with File ID: BF141326.D met requirements for all samples except for 4-Chloroaniline[29%], Hexachlorocyclopentadiene[160%], marginally low therefore no corrective action was taken.

The Blank Spike Duplicate for {PB166363BSD} with File ID: BF141327.D met requirements for all samples except for 3-Nitroaniline[40%], 4-Chloroaniline[25%] and Hexachlorocyclopentadiene[160%],marginally low therefore no corrective action was taken.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

E. Additional Comments:

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

CASE NARRATIVE

Weston Solutions

Project Name: Ft Meade Tipton Airfield Parcel RI - PO 0111169

Project # N/A

Chemtech Project # Q1211

Test Name: PESTICIDE Group1

A. Number of Samples and Date of Receipt:

3 Water samples were received on 01/29/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Ammonia, Anions Group5, Diesel Range Organics, Gasoline Range Organics, Hardness, Total, Hexavalent Chromium, Mercury, Metals ICP-TAL, METALS-TAL, Oil and Grease, PESTICIDE Group1, SVOC-TCL BNA -20, TOC and VOC-TCLVOA-10. This data package contains results for PESTICIDE Group1.

C. Analytical Techniques:

The analysis was performed on instrument ECD_L. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalog # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HMG017- 11. The analysis of PESTICIDE Group1s was based on method 8081B and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for TAPHHA-MW01-012825-00-T4 [Tetrachloro-m-xylene(1) - 125%].as per method one surrogate is allowed to failed, therefore no corrective action was taken.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

E. Calculation for Concentration in Water Samples:

$$\text{Concentration in ug/L} = \frac{(Ax) (Vt) (DF) (GPC)}{(CF) (Vo) (Vi)}$$

Where,

Ax = Response (peak area or height) of the compound to be measured.

CF = Mean Calibration Factor from the initial calibration (area/ng).

Vo = Volume of water extracted in mL.

Vi = Volume of extract injected in uL.

Vt = Volume of the concentrated extract in uL

GPC = $\frac{V_{in}}{V_{out}}$ = GPC factor (If no GPC is performed, GPC=1)

Vin = Volume of extract loaded onto GPC column.

Vout = Volume of extract collected after GPC cleanup.

DF = Dilution Factor.

F. Additional Comments:

The not QT review data is reported in the Miscellaneous

G. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

CASE NARRATIVE

Weston Solutions

Project Name: Ft Meade Tipton Airfield Parcel RI - PO 0111169

Project # N/A

Chemtech Project # Q1211

Test Name: Diesel Range Organics

A. Number of Samples and Date of Receipt:

3 Water samples were received on 01/29/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Ammonia, Anions Group5, Diesel Range Organics, Gasoline Range Organics, Hardness, Total, Hexavalent Chromium, Mercury, Metals ICP-TAL, METALS-TAL, Oil and Grease, PESTICIDE Group1, SVOC-TCL BNA -20, TOC and VOC-TCLVOA-10. This data package contains results for Diesel Range Organics.

C. Analytical Techniques:

The analysis were performed on instrument FID_E. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 um df, catalog 10224. The analysis of Diesel Range Organics was based on method 8015D and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

E. Additional Comments:

The not QT review data is reported in the Miscellaneous.

F. Calculation for Concentration in WATER samples :

The sample concentrations (Cs) in ug/L are calculated as follows:

$$Cs = \frac{\{\text{Extract DRO-net (ug/mL)}\} \{\text{Final vol. extract (mL)}\} \{\text{Df}\}}{Ws}$$

Where

DRO (net)ug/mL = DRO (total) ug /mL - DRO (solvent) ug /mL

Df = Dilution factor

Ws= Weight of sample in mL

G. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

CASE NARRATIVE

Weston Solutions

Project Name: Ft Meade Tipton Airfield Parcel RI - PO 0111169

Project # N/A

Chemtech Project # Q1211

Test Name: Metals ICP-TAL,Mercury

A. Number of Samples and Date of Receipt:

3 Water samples were received on 01/29/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Ammonia, Anions Group5, Diesel Range Organics, Gasoline Range Organics, Hardness, Total, Hexavalent Chromium, Mercury, Metals ICP-TAL, METALS-TAL, Oil and Grease, PESTICIDE Group1, SVOC-TCL BNA -20, TOC and VOC-TCLVOA-10. This data package contains results for Metals ICP-TAL,Mercury.

C. Analytical Techniques:

The analysis of Metals ICP-TAL was based on method 6020B, digestion based on method 3010 (waters). The analysis and digestion of Mercury was based on method 7470A.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate (TAPIAL2-MW03-012825-00-T3DUP) analysis met criteria for all samples except for Aluminum due to matrix interference.

The Matrix Spike (TAPIAL2-MW03-012825-00-T3MS) analysis met criteria for all samples except for Aluminum, Arsenic, Calcium, Iron, Potassium, Silver due to matrix interference.

The Matrix Spike Duplicate (TAPHHA-MW12-012725-00-T2MSD) analysis met criteria for all samples except for Mercury due to matrix interference.

The Matrix Spike Duplicate (TAPIAL2-MW03-012825-00-T3MSD) analysis met criteria for all samples except for Aluminum, Arsenic, Calcium, Iron, Potassium, Silver due to matrix interference.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met criteria for all samples.

E. Calculations:

Calculation for ICP-MS Water Sample:

$$\text{Concentration or Result } (\mu\text{g/L}) = C \times \frac{V_f}{V_i} \times \text{DF}$$

Where,

C = Instrument value in ppb (The average of all replicate integrations)

Vf = Final digestion volume (mL)

Vi = Initial aliquot amount (mL) (Sample amount taken in prep)

DF = Dilution Factor

Calculation for Hg Water Sample:

$$\text{Concentration or Result } (\mu\text{g/L}) = C \times \text{DF}$$

Where,

C = Instrument response in $\mu\text{g/L}$ from the calibration curve.

DF = Dilution Factor

F. Additional Comments:

In analytical sequence LB134616, The % recovery was outside of acceptance limit for Nickel of LLCCV, Recovery failing marginally so no corrective action was taken.

Collision cell is being used to remove potential interferences. The analytes Na, Mg, Al, K, V, Cr, Mn, Fe, Co, Ni, Cu, Zn, As are being analyzed with collision cell and analytes Be, B, Ca, Ti, Se, Sr, Zr, Mo, Ag, Cd, Sn, Sb, Ba, Tl, Pb, U are being analyzed with Non-Collision Cell. Helium gas is used for the Collision Cell analysis.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

CASE NARRATIVE

Weston Solutions

Project Name: Ft Meade Tipton Airfield Parcel RI - PO 0111169

Project # N/A

Chemtech Project # Q1211

Test Name: Hexavalent Chromium, Oil and Grease, Anions Group5, TOC, Ammonia

A. Number of Samples and Date of Receipt:

3 Water samples were received on 01/29/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Ammonia, Anions Group5, Diesel Range Organics, Gasoline Range Organics, Hardness, Total, Hexavalent Chromium, Mercury, Metals ICP-TAL, METALS-TAL, Oil and Grease, PESTICIDE Group1, SVOC-TCL BNA -20, TOC and VOC-TCLVOA-10. This data package contains results for Hexavalent Chromium, Oil and Grease, Anions Group5, TOC, Ammonia.

C. Analytical Techniques:

The analysis of Oil and Grease was based on method 1664A, The analysis of Hexavalent Chromium was based on method 7196A, The analysis of Anions Group5 was based on method 9056A, The analysis of TOC was based on method 9060A and The analysis of Ammonia was based on method SM4500-NH3.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

Sample TAPHHA-MW01-012825-00-T4 was diluted due to high concentrations for Chloride.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike analysis met criteria for all samples.

The Matrix Spike Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

As per method 1664A, MS/MSD is required to be performed with the sample analysis. However, Lab did not receive sufficient volume to perform the MS/MSD for Q1211, therefore Lab reported MS-MSD from Q1252.



I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
OR	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1211

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 02/11/2025

Hit Summary Sheet SW-846

SDG No.: Q1211
Client: Weston Solutions

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID:	TAPIAL2-MW03-012825-00-T3								
Q1211-02	TAPIAL2-MW03-0	Water	Chloromethane	0.49	J	0.35	0.50	5.00	ug/L
			Total Voc :	0.49					
			Total Concentration:	0.49					

A

B

C

D



SAMPLE DATA

Report of Analysis

Client:	Weston Solutions		Date Collected:	01/28/25	
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169		Date Received:	01/29/25	
Client Sample ID:	TAPHHA-MW01-012825-00-T4		SDG No.:	Q1211	
Lab Sample ID:	Q1211-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085583.D	1		01/30/25 16:42	VN013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	5.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	5.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	5.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.50	U	0.56	0.50	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	5.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	25.0	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	5.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	5.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	5.00	ug/L
110-82-7	Cyclohexane	3.80	U	1.60	3.80	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.50	U	0.25	0.50	5.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	5.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.75	U	0.19	0.75	5.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	5.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.75	U	0.24	0.75	5.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	5.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	25.0	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	5.00	ug/L

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPHHA-MW01-012825-00-T4	SDG No.:	Q1211
Lab Sample ID:	Q1211-01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085583.D	1		01/30/25 16:42	VN013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	5.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	25.0	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	5.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	5.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	5.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	5.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	10.0	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	5.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	5.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	5.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	2.00	U	0.46	2.00	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.50	U	0.51	0.50	5.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	57.5		81 - 118		115%	SPK: 50
1868-53-7	Dibromofluoromethane	54.0		80 - 119		108%	SPK: 50
2037-26-5	Toluene-d8	49.6		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	44.2		85 - 114		88%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	161000	8.224				
540-36-3	1,4-Difluorobenzene	307000	9.1				
3114-55-4	Chlorobenzene-d5	272000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	106000	13.788				

Report of Analysis

Client:	Weston Solutions		Date Collected:	01/28/25	
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169		Date Received:	01/29/25	
Client Sample ID:	TAPHHA-MW01-012825-00-T4		SDG No.:	Q1211	
Lab Sample ID:	Q1211-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085583.D	1		01/30/25 16:42	VN013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPIAL2-MW03-012825-00-T3	SDG No.:	Q1211
Lab Sample ID:	Q1211-02	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085584.D	1		01/30/25 17:06	VN013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	5.00	ug/L
74-87-3	Chloromethane	0.49	J	0.35	0.50	5.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	5.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.50	U	0.56	0.50	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	5.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	25.0	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	5.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	5.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	5.00	ug/L
110-82-7	Cyclohexane	3.80	U	1.60	3.80	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.50	U	0.25	0.50	5.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	5.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.75	U	0.19	0.75	5.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	5.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.75	U	0.24	0.75	5.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	5.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	25.0	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	5.00	ug/L

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPIAL2-MW03-012825-00-T3	SDG No.:	Q1211
Lab Sample ID:	Q1211-02	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085584.D	1		01/30/25 17:06	VN013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	5.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	25.0	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	5.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	5.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	5.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	5.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	10.0	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	5.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	5.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	5.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	2.00	U	0.46	2.00	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.50	U	0.51	0.50	5.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.5		81 - 118		111%	SPK: 50
1868-53-7	Dibromofluoromethane	52.1		80 - 119		104%	SPK: 50
2037-26-5	Toluene-d8	49.2		89 - 112		98%	SPK: 50
460-00-4	4-Bromofluorobenzene	41.4	*	85 - 114		83%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	173000	8.224				
540-36-3	1,4-Difluorobenzene	322000	9.1				
3114-55-4	Chlorobenzene-d5	276000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	104000	13.794				

Report of Analysis

Client:	Weston Solutions		Date Collected:	01/28/25	
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169		Date Received:	01/29/25	
Client Sample ID:	TAPIAL2-MW03-012825-00-T3		SDG No.:	Q1211	
Lab Sample ID:	Q1211-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085584.D	1		01/30/25 17:06	VN013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Weston Solutions		Date Collected:	01/28/25	
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169		Date Received:	01/29/25	
Client Sample ID:	TAPIAL2-MW03-012825-00-T3RE		SDG No.:	Q1211	
Lab Sample ID:	Q1211-02RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085595.D	1		01/31/25 12:03	VN013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	5.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	5.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	5.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.50	U	0.56	0.50	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	5.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	25.0	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	5.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	5.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	5.00	ug/L
110-82-7	Cyclohexane	3.80	U	1.60	3.80	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.50	U	0.25	0.50	5.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	5.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.75	U	0.19	0.75	5.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	5.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.75	U	0.24	0.75	5.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	5.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	25.0	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	5.00	ug/L

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPIAL2-MW03-012825-00-T3RE	SDG No.:	Q1211
Lab Sample ID:	Q1211-02RE	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085595.D	1		01/31/25 12:03	VN013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	5.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	25.0	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	5.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	5.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	5.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	5.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	10.0	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	5.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	5.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	5.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	2.00	U	0.46	2.00	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.50	U	0.51	0.50	5.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	56.7		81 - 118		113%	SPK: 50
1868-53-7	Dibromofluoromethane	52.8		80 - 119		106%	SPK: 50
2037-26-5	Toluene-d8	49.2		89 - 112		98%	SPK: 50
460-00-4	4-Bromofluorobenzene	45.8		85 - 114		92%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	166000	8.224				
540-36-3	1,4-Difluorobenzene	310000	9.1				
3114-55-4	Chlorobenzene-d5	268000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	102000	13.788				

Report of Analysis

Client:	Weston Solutions		Date Collected:	01/28/25	
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169		Date Received:	01/29/25	
Client Sample ID:	TAPIAL2-MW03-012825-00-T3RE		SDG No.:	Q1211	
Lab Sample ID:	Q1211-02RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085595.D	1		01/31/25 12:03	VN013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAP-TB-02-012825-T4	SDG No.:	Q1211
Lab Sample ID:	Q1211-03	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085582.D	1		01/30/25 16:18	VN013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	5.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	5.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	5.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.50	U	0.56	0.50	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	5.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	25.0	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	5.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	5.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	5.00	ug/L
110-82-7	Cyclohexane	3.80	U	1.60	3.80	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.50	U	0.25	0.50	5.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	5.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.75	U	0.19	0.75	5.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	5.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.75	U	0.24	0.75	5.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	5.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	25.0	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	5.00	ug/L

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAP-TB-02-012825-T4	SDG No.:	Q1211
Lab Sample ID:	Q1211-03	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085582.D	1		01/30/25 16:18	VN013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	5.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	25.0	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	5.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	5.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	5.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	5.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	10.0	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	5.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	5.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	5.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	2.00	U	0.46	2.00	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.50	U	0.51	0.50	5.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	56.9		81 - 118		114%	SPK: 50
1868-53-7	Dibromofluoromethane	53.3		80 - 119		107%	SPK: 50
2037-26-5	Toluene-d8	50.1		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	42.9		85 - 114		86%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	164000	8.224				
540-36-3	1,4-Difluorobenzene	307000	9.1				
3114-55-4	Chlorobenzene-d5	275000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	105000	13.788				

Report of Analysis

Client:	Weston Solutions		Date Collected:	01/28/25	
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169		Date Received:	01/29/25	
Client Sample ID:	TAP-TB-02-012825-T4		SDG No.:	Q1211	
Lab Sample ID:	Q1211-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085582.D	1		01/30/25 16:18	VN013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q1211	OrderDate:	1/29/2025 10:10:00 AM
Client:	Weston Solutions	Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169
Contact:	Nathan Fretz	Location:	N31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1211-01	TAPHHA-MW01-01282 5-00-T4	Water			01/28/25			01/29/25
			VOC-TCLVOA-10	8260D			01/30/25	
Q1211-02	TAPIAL2-MW03-0128 25-00-T3	Water			01/28/25			01/29/25
			VOC-TCLVOA-10	8260D			01/30/25	
Q1211-02RE	TAPIAL2-MW03-0128 25-00-T3RE	Water			01/28/25			01/29/25
			VOC-TCLVOA-10	8260D			01/31/25	
Q1211-03	TAP-TB-02-012825-T4	Water			01/28/25			01/29/25
			VOC-TCLVOA-10	8260D			01/30/25	



SAMPLE DATA

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPHHA-MW01-012825-00-T4	SDG No.:	Q1211
Lab Sample ID:	Q1211-01	Matrix:	Water
Analytical Method:	8015D GRO	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Gasoline Range Organics
GPC Factor :		Injection Volume :	
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
FB031377.D	1	01/29/25 19:01	FB012925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
GRO	GRO	9.00	U	6.00	9.00	45.0	ug/L
SURROGATES							
98-08-8	Alpha,Alpha,Alpha-Trifluoroto	21.6		50 - 150		108%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPIAL2-MW03-012825-00-T3	SDG No.:	Q1211
Lab Sample ID:	Q1211-02	Matrix:	Water
Analytical Method:	8015D GRO	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Gasoline Range Organics
GPC Factor :		Injection Volume :	
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
FB031378.D	1	01/29/25 19:28	FB012925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
GRO	GRO	10.0	J	6.00	9.00	45.0	ug/L
SURROGATES							
98-08-8	Alpha,Alpha,Alpha-Trifluoroto	18.9		50 - 150		95%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	Q1211	OrderDate:	1/29/2025 10:10:00 AM
Client:	Weston Solutions	Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169
Contact:	Nathan Fretz	Location:	N31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1211-01	TAPHHA-MW01-01282 5-00-T4	Water			01/28/25			01/29/25
			Diesel Range Organics	8015D		01/30/25	01/30/25	
			Gasoline Range Organics	8015D			01/29/25	
Q1211-02	TAPIAL2-MW03-0128 25-00-T3	Water			01/28/25			01/29/25
			Diesel Range Organics	8015D		01/30/25	01/30/25	
			Gasoline Range Organics	8015D			01/29/25	

Hit Summary Sheet SW-846

SDG No.: Q1211
Client: Weston Solutions

Sample ID	Client ID	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID : TAPHHA-MW01-012825-00-T4								
Q1211-01	TAPHHA-MW01-012825 WATER	2-Pentanone, 4-hydroxy-4-methyl *	3.400	A	0		0	ug/L
Q1211-01	TAPHHA-MW01-012825 WATER	Palmitoleic acid *	2.600	J	0		0	ug/L
Total Tics :				6.00				
Total Concentration:				6.00				
Client ID : TAPIAL2-MW03-012825-00-T3								
Q1211-02	TAPIAL2-MW03-012825 WATER	Benzophenone *	2.600	J	0		0	ug/L
Q1211-02	TAPIAL2-MW03-012825 WATER	n-Hexadecanoic acid *	2.600	J	0		0	ug/L
Q1211-02	TAPIAL2-MW03-012825 WATER	Propane, 1,1-dimethoxy- *	2.100	JB	0		0	ug/L
Total Tics :				7.30				
Total Concentration:				7.30				



SAMPLE DATA

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPHHA-MW01-012825-00-T4	SDG No.:	Q1211
Lab Sample ID:	Q1211-01	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-TCL BNA -20
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141328.D	1	01/30/25 08:15	01/31/25 13:27	PB166363

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
100-52-7	Benzaldehyde	8.00	U	4.00	8.00	10.0	ug/L
108-95-2	Phenol	4.00	U	0.93	4.00	5.00	ug/L
111-44-4	bis(2-Chloroethyl)ether	4.00	U	1.20	4.00	5.00	ug/L
95-57-8	2-Chlorophenol	4.00	U	0.71	4.00	5.00	ug/L
95-48-7	2-Methylphenol	4.00	U	1.10	4.00	5.00	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	4.00	U	1.40	4.00	5.00	ug/L
98-86-2	Acetophenone	4.00	U	1.10	4.00	5.00	ug/L
65794-96-9	3+4-Methylphenols	8.00	U	1.20	8.00	10.0	ug/L
621-64-7	n-Nitroso-di-n-propylamine	2.50	U	1.50	2.50	2.50	ug/L
67-72-1	Hexachloroethane	4.00	U	1.00	4.00	5.00	ug/L
98-95-3	Nitrobenzene	4.00	U	1.30	4.00	5.00	ug/L
78-59-1	Isophorone	4.00	U	1.10	4.00	5.00	ug/L
88-75-5	2-Nitrophenol	4.00	U	2.00	4.00	5.00	ug/L
105-67-9	2,4-Dimethylphenol	4.00	U	1.50	4.00	5.00	ug/L
111-91-1	bis(2-Chloroethoxy)methane	4.00	U	1.00	4.00	5.00	ug/L
120-83-2	2,4-Dichlorophenol	4.00	U	0.88	4.00	5.00	ug/L
91-20-3	Naphthalene	4.00	U	1.00	4.00	5.00	ug/L
106-47-8	4-Chloroaniline	4.00	UQ	1.30	4.00	5.00	ug/L
87-68-3	Hexachlorobutadiene	4.00	U	1.30	4.00	5.00	ug/L
105-60-2	Caprolactam	8.00	U	1.70	8.00	10.0	ug/L
59-50-7	4-Chloro-3-methylphenol	4.00	U	0.84	4.00	5.00	ug/L
91-57-6	2-Methylnaphthalene	4.00	U	1.10	4.00	5.00	ug/L
77-47-4	Hexachlorocyclopentadiene	8.00	UQ	5.00	8.00	10.0	ug/L
88-06-2	2,4,6-Trichlorophenol	4.00	U	0.89	4.00	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	4.00	U	1.00	4.00	5.00	ug/L
92-52-4	1,1-Biphenyl	4.00	U	0.91	4.00	5.00	ug/L
91-58-7	2-Chloronaphthalene	4.00	U	0.97	4.00	5.00	ug/L
88-74-4	2-Nitroaniline	4.00	U	1.40	4.00	5.00	ug/L
131-11-3	Dimethylphthalate	4.00	U	0.93	4.00	5.00	ug/L

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPHHA-MW01-012825-00-T4	SDG No.:	Q1211
Lab Sample ID:	Q1211-01	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-TCL BNA -20
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141328.D	1	01/30/25 08:15	01/31/25 13:27	PB166363

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
208-96-8	Acenaphthylene	4.00	U	1.00	4.00	5.00	ug/L
606-20-2	2,6-Dinitrotoluene	4.00	U	1.20	4.00	5.00	ug/L
99-09-2	3-Nitroaniline	4.00	UQ	1.40	4.00	5.00	ug/L
83-32-9	Acenaphthene	4.00	U	0.81	4.00	5.00	ug/L
51-28-5	2,4-Dinitrophenol	8.00	U	6.40	8.00	10.0	ug/L
100-02-7	4-Nitrophenol	8.00	U	2.00	8.00	10.0	ug/L
132-64-9	Dibenzofuran	4.00	U	0.93	4.00	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	4.00	U	1.50	4.00	5.00	ug/L
84-66-2	Diethylphthalate	4.00	U	1.00	4.00	5.00	ug/L
7005-72-3	4-Chlorophenyl-phenylether	4.00	U	0.98	4.00	5.00	ug/L
86-73-7	Fluorene	4.00	U	0.96	4.00	5.00	ug/L
100-01-6	4-Nitroaniline	4.00	U	2.00	4.00	5.00	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	8.00	U	3.10	8.00	10.0	ug/L
86-30-6	n-Nitrosodiphenylamine	4.00	U	0.89	4.00	5.00	ug/L
101-55-3	4-Bromophenyl-phenylether	4.00	U	0.95	4.00	5.00	ug/L
118-74-1	Hexachlorobenzene	4.00	U	1.10	4.00	5.00	ug/L
1912-24-9	Atrazine	4.00	U	1.30	4.00	5.00	ug/L
87-86-5	Pentachlorophenol	8.00	U	1.90	8.00	10.0	ug/L
85-01-8	Phenanthrene	4.00	U	0.89	4.00	5.00	ug/L
120-12-7	Anthracene	4.00	U	1.10	4.00	5.00	ug/L
86-74-8	Carbazole	4.00	U	1.20	4.00	5.00	ug/L
84-74-2	Di-n-butylphthalate	4.00	U	1.50	4.00	5.00	ug/L
206-44-0	Fluoranthene	4.00	U	1.30	4.00	5.00	ug/L
129-00-0	Pyrene	4.00	U	1.10	4.00	5.00	ug/L
85-68-7	Butylbenzylphthalate	4.00	U	2.10	4.00	5.00	ug/L
91-94-1	3,3-Dichlorobenzidine	8.00	U	1.30	8.00	10.0	ug/L
56-55-3	Benzo(a)anthracene	4.00	U	0.94	4.00	5.00	ug/L
218-01-9	Chrysene	4.00	U	0.86	4.00	5.00	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	4.00	U	1.90	4.00	5.00	ug/L
117-84-0	Di-n-octyl phthalate	8.00	U	2.50	8.00	10.0	ug/L
205-99-2	Benzo(b)fluoranthene	4.00	U	1.10	4.00	5.00	ug/L

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPHHA-MW01-012825-00-T4	SDG No.:	Q1211
Lab Sample ID:	Q1211-01	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-TCL BNA -20
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141328.D	1	01/30/25 08:15	01/31/25 13:27	PB166363

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	4.00	U	1.20	4.00	5.00	ug/L
50-32-8	Benzo(a)pyrene	4.00	U	1.70	4.00	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	4.00	U	1.00	4.00	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	4.00	U	1.20	4.00	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	4.00	U	1.20	4.00	5.00	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	4.00	U	1.10	4.00	5.00	ug/L
123-91-1	1,4-Dioxane	4.00	U	1.30	4.00	5.00	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	4.00	U	0.79	4.00	5.00	ug/L
SURROGATES							
367-12-4	2-Fluorophenol	57.4		19 - 119		38%	SPK: 150
13127-88-3	Phenol-d6	35.8		10 - 130		24%	SPK: 150
4165-60-0	Nitrobenzene-d5	85.4		44 - 120		85%	SPK: 100
321-60-8	2-Fluorobiphenyl	80.4		44 - 119		80%	SPK: 100
118-79-6	2,4,6-Tribromophenol	135		43 - 140		90%	SPK: 150
1718-51-0	Terphenyl-d14	68.7		50 - 134		69%	SPK: 100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	155000	6.798				
1146-65-2	Naphthalene-d8	631000	8.081				
15067-26-2	Acenaphthene-d10	343000	9.833				
1517-22-2	Phenanthrene-d10	611000	11.322				
1719-03-5	Chrysene-d12	458000	13.963				
1520-96-3	Perylene-d12	357000	15.427				
TENTATIVE IDENTIFIED COMPOUNDS							
000123-42-2	2-Pentanone, 4-hydroxy-4-methyl-	3.40	A			5.00	ug/L
000373-49-9	Palmitoleic acid	2.60	J			12.6	ug/L

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPHHA-MW01-012825-00-T4	SDG No.:	Q1211
Lab Sample ID:	Q1211-01	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-TCL BNA -20
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141328.D	1	01/30/25 08:15	01/31/25 13:27	PB166363

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPIAL2-MW03-012825-00-T3	SDG No.:	Q1211
Lab Sample ID:	Q1211-02	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-TCL BNA -20
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141329.D	1	01/30/25 08:15	01/31/25 13:53	PB166363

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
100-52-7	Benzaldehyde	8.00	U	4.00	8.00	10.0	ug/L
108-95-2	Phenol	4.00	U	0.93	4.00	5.00	ug/L
111-44-4	bis(2-Chloroethyl)ether	4.00	U	1.20	4.00	5.00	ug/L
95-57-8	2-Chlorophenol	4.00	U	0.71	4.00	5.00	ug/L
95-48-7	2-Methylphenol	4.00	U	1.10	4.00	5.00	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	4.00	U	1.40	4.00	5.00	ug/L
98-86-2	Acetophenone	4.00	U	1.10	4.00	5.00	ug/L
65794-96-9	3+4-Methylphenols	8.00	U	1.20	8.00	10.0	ug/L
621-64-7	n-Nitroso-di-n-propylamine	2.50	U	1.50	2.50	2.50	ug/L
67-72-1	Hexachloroethane	4.00	U	1.00	4.00	5.00	ug/L
98-95-3	Nitrobenzene	4.00	U	1.30	4.00	5.00	ug/L
78-59-1	Isophorone	4.00	U	1.10	4.00	5.00	ug/L
88-75-5	2-Nitrophenol	4.00	U	2.00	4.00	5.00	ug/L
105-67-9	2,4-Dimethylphenol	4.00	U	1.50	4.00	5.00	ug/L
111-91-1	bis(2-Chloroethoxy)methane	4.00	U	1.00	4.00	5.00	ug/L
120-83-2	2,4-Dichlorophenol	4.00	U	0.88	4.00	5.00	ug/L
91-20-3	Naphthalene	4.00	U	1.00	4.00	5.00	ug/L
106-47-8	4-Chloroaniline	4.00	UQ	1.30	4.00	5.00	ug/L
87-68-3	Hexachlorobutadiene	4.00	U	1.30	4.00	5.00	ug/L
105-60-2	Caprolactam	8.00	U	1.70	8.00	10.0	ug/L
59-50-7	4-Chloro-3-methylphenol	4.00	U	0.84	4.00	5.00	ug/L
91-57-6	2-Methylnaphthalene	4.00	U	1.10	4.00	5.00	ug/L
77-47-4	Hexachlorocyclopentadiene	8.00	UQ	5.00	8.00	10.0	ug/L
88-06-2	2,4,6-Trichlorophenol	4.00	U	0.89	4.00	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	4.00	U	1.00	4.00	5.00	ug/L
92-52-4	1,1-Biphenyl	4.00	U	0.91	4.00	5.00	ug/L
91-58-7	2-Chloronaphthalene	4.00	U	0.97	4.00	5.00	ug/L
88-74-4	2-Nitroaniline	4.00	U	1.40	4.00	5.00	ug/L
131-11-3	Dimethylphthalate	4.00	U	0.93	4.00	5.00	ug/L

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPIAL2-MW03-012825-00-T3	SDG No.:	Q1211
Lab Sample ID:	Q1211-02	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-TCL BNA -20
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141329.D	1	01/30/25 08:15	01/31/25 13:53	PB166363

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
208-96-8	Acenaphthylene	4.00	U	1.00	4.00	5.00	ug/L
606-20-2	2,6-Dinitrotoluene	4.00	U	1.20	4.00	5.00	ug/L
99-09-2	3-Nitroaniline	4.00	UQ	1.40	4.00	5.00	ug/L
83-32-9	Acenaphthene	4.00	U	0.81	4.00	5.00	ug/L
51-28-5	2,4-Dinitrophenol	8.00	U	6.40	8.00	10.0	ug/L
100-02-7	4-Nitrophenol	8.00	U	2.00	8.00	10.0	ug/L
132-64-9	Dibenzofuran	4.00	U	0.93	4.00	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	4.00	U	1.50	4.00	5.00	ug/L
84-66-2	Diethylphthalate	4.00	U	1.00	4.00	5.00	ug/L
7005-72-3	4-Chlorophenyl-phenylether	4.00	U	0.98	4.00	5.00	ug/L
86-73-7	Fluorene	4.00	U	0.96	4.00	5.00	ug/L
100-01-6	4-Nitroaniline	4.00	U	2.00	4.00	5.00	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	8.00	U	3.10	8.00	10.0	ug/L
86-30-6	n-Nitrosodiphenylamine	4.00	U	0.89	4.00	5.00	ug/L
101-55-3	4-Bromophenyl-phenylether	4.00	U	0.95	4.00	5.00	ug/L
118-74-1	Hexachlorobenzene	4.00	U	1.10	4.00	5.00	ug/L
1912-24-9	Atrazine	4.00	U	1.30	4.00	5.00	ug/L
87-86-5	Pentachlorophenol	8.00	U	1.90	8.00	10.0	ug/L
85-01-8	Phenanthrene	4.00	U	0.89	4.00	5.00	ug/L
120-12-7	Anthracene	4.00	U	1.10	4.00	5.00	ug/L
86-74-8	Carbazole	4.00	U	1.20	4.00	5.00	ug/L
84-74-2	Di-n-butylphthalate	4.00	U	1.50	4.00	5.00	ug/L
206-44-0	Fluoranthene	4.00	U	1.30	4.00	5.00	ug/L
129-00-0	Pyrene	4.00	U	1.10	4.00	5.00	ug/L
85-68-7	Butylbenzylphthalate	4.00	U	2.10	4.00	5.00	ug/L
91-94-1	3,3-Dichlorobenzidine	8.00	U	1.30	8.00	10.0	ug/L
56-55-3	Benzo(a)anthracene	4.00	U	0.94	4.00	5.00	ug/L
218-01-9	Chrysene	4.00	U	0.86	4.00	5.00	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	4.00	U	1.90	4.00	5.00	ug/L
117-84-0	Di-n-octyl phthalate	8.00	U	2.50	8.00	10.0	ug/L
205-99-2	Benzo(b)fluoranthene	4.00	U	1.10	4.00	5.00	ug/L

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPIAL2-MW03-012825-00-T3	SDG No.:	Q1211
Lab Sample ID:	Q1211-02	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-TCL BNA -20
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141329.D	1	01/30/25 08:15	01/31/25 13:53	PB166363

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	4.00	U	1.20	4.00	5.00	ug/L
50-32-8	Benzo(a)pyrene	4.00	U	1.70	4.00	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	4.00	U	1.00	4.00	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	4.00	U	1.20	4.00	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	4.00	U	1.20	4.00	5.00	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	4.00	U	1.10	4.00	5.00	ug/L
123-91-1	1,4-Dioxane	4.00	U	1.30	4.00	5.00	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	4.00	U	0.79	4.00	5.00	ug/L
SURROGATES							
367-12-4	2-Fluorophenol	57.1		19 - 119		38%	SPK: 150
13127-88-3	Phenol-d6	35.2		10 - 130		23%	SPK: 150
4165-60-0	Nitrobenzene-d5	87.1		44 - 120		87%	SPK: 100
321-60-8	2-Fluorobiphenyl	81.2		44 - 119		81%	SPK: 100
118-79-6	2,4,6-Tribromophenol	134		43 - 140		89%	SPK: 150
1718-51-0	Terphenyl-d14	64.9		50 - 134		65%	SPK: 100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	150000	6.798				
1146-65-2	Naphthalene-d8	599000	8.081				
15067-26-2	Acenaphthene-d10	330000	9.834				
1517-22-2	Phenanthrene-d10	567000	11.322				
1719-03-5	Chrysene-d12	401000	13.963				
1520-96-3	Perylene-d12	320000	15.421				
TENTATIVE IDENTIFIED COMPOUNDS							
004744-10-9	Propane, 1,1-dimethoxy-	2.10	JB			2.15	ug/L
000119-61-9	Benzophenone	2.60	J			10.6	ug/L
000057-10-3	n-Hexadecanoic acid	2.60	J			11.9	ug/L

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPIAL2-MW03-012825-00-T3	SDG No.:	Q1211
Lab Sample ID:	Q1211-02	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-TCL BNA -20
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141329.D	1	01/30/25 08:15	01/31/25 13:53	PB166363

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q1211	OrderDate:	1/29/2025 10:10:00 AM
Client:	Weston Solutions	Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169
Contact:	Nathan Fretz	Location:	N31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1211-01	TAPHHA-MW01-01282 5-00-T4	Water	SVOC-TCL BNA -20	8270E	01/28/25	01/30/25	01/31/25	01/29/25
Q1211-02	TAPIAL2-MW03-0128 25-00-T3	Water	SVOC-TCL BNA -20	8270E	01/28/25	01/30/25	01/31/25	01/29/25

Hit Summary Sheet
SW-846

A

B

C

D

SDG No.:	Q1211	Order ID:	Q1211
Client:	Weston Solutions	Project ID:	Ft Meade Tipton Airfield Parcel RI - P

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID :				Total Concentration:	0.000				



SAMPLE DATA

Report of Analysis

Client:	Weston Solutions		Date Collected:	01/28/25	
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169		Date Received:	01/29/25	
Client Sample ID:	TAPHHA-MW01-012825-00-T4		SDG No.:	Q1211	
Lab Sample ID:	Q1211-01		Matrix:	WATER	
Analytical Method:	SW8081		% Solid:	0	Decanted:
Sample Wt/Vol:	990	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PESTICIDE Group1	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093938.D	1	01/30/25 09:25	01/31/25 14:02	PB166365

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
319-84-6	alpha-BHC	0.025	U	0.0062	0.025	0.051	ug/L
319-85-7	beta-BHC	0.025	U	0.014	0.025	0.051	ug/L
319-86-8	delta-BHC	0.025	U	0.015	0.025	0.051	ug/L
58-89-9	gamma-BHC (Lindane)	0.025	U	0.0049	0.025	0.051	ug/L
76-44-8	Heptachlor	0.025	U	0.0055	0.025	0.051	ug/L
309-00-2	Aldrin	0.025	U	0.0044	0.025	0.051	ug/L
1024-57-3	Heptachlor epoxide	0.025	U	0.0091	0.025	0.051	ug/L
959-98-8	Endosulfan I	0.025	U	0.0051	0.025	0.051	ug/L
60-57-1	Dieldrin	0.025	U	0.0047	0.025	0.051	ug/L
72-55-9	4,4-DDE	0.025	U	0.0045	0.025	0.051	ug/L
72-20-8	Endrin	0.010	U	0.0043	0.010	0.051	ug/L
33213-65-9	Endosulfan II	0.025	U	0.0076	0.025	0.051	ug/L
72-54-8	4,4-DDD	0.025	U	0.0093	0.025	0.051	ug/L
1031-07-8	Endosulfan Sulfate	0.025	U	0.0035	0.025	0.051	ug/L
50-29-3	4,4-DDT	0.025	U	0.0044	0.025	0.051	ug/L
72-43-5	Methoxychlor	0.025	U	0.011	0.025	0.051	ug/L
53494-70-5	Endrin ketone	0.025	U	0.0098	0.025	0.051	ug/L
7421-93-4	Endrin aldehyde	0.025	U	0.010	0.025	0.051	ug/L
5103-71-9	alpha-Chlordane	0.025	U	0.0061	0.025	0.051	ug/L
5103-74-2	gamma-Chlordane	0.025	U	0.0061	0.025	0.051	ug/L
8001-35-2	Toxaphene	0.51	U	0.15	0.51	1.00	ug/L
57-74-9	Chlordane	0.25	U	0.083	0.25	0.51	ug/L
2385-85-5	Mirex	0.025	U	0.0041	0.025	0.051	ug/L
SURROGATES							
2051-24-3	Decachlorobiphenyl	20.1		30 - 135		100%	SPK: 20
877-09-8	Tetrachloro-m-xylene	25.1	*	44 - 124		125%	SPK: 20

Report of Analysis

Client:	Weston Solutions			Date Collected:	01/28/25	
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169			Date Received:	01/29/25	
Client Sample ID:	TAPHHA-MW01-012825-00-T4			SDG No.:	Q1211	
Lab Sample ID:	Q1211-01			Matrix:	WATER	
Analytical Method:	SW8081			% Solid:	0	Decanted:
Sample Wt/Vol:	990	Units:	mL	Final Vol:	10000	uL
Soil Aliquot Vol:			uL	Test:	PESTICIDE Group1	
Extraction Type:				Injection Volume :		
GPC Factor :	1.0	PH :				
Prep Method :	3510C					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093938.D	1	01/30/25 09:25	01/31/25 14:02	PB166365

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions		Date Collected:	01/28/25	
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169		Date Received:	01/29/25	
Client Sample ID:	TAPIAL2-MW03-012825-00-T3		SDG No.:	Q1211	
Lab Sample ID:	Q1211-02		Matrix:	WATER	
Analytical Method:	SW8081		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PESTICIDE Group1	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093939.D	1	01/30/25 09:25	01/31/25 14:15	PB166365

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
319-84-6	alpha-BHC	0.025	U	0.0061	0.025	0.050	ug/L
319-85-7	beta-BHC	0.025	U	0.014	0.025	0.050	ug/L
319-86-8	delta-BHC	0.025	U	0.015	0.025	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.025	U	0.0049	0.025	0.050	ug/L
76-44-8	Heptachlor	0.025	U	0.0054	0.025	0.050	ug/L
309-00-2	Aldrin	0.025	U	0.0044	0.025	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.025	U	0.0090	0.025	0.050	ug/L
959-98-8	Endosulfan I	0.025	U	0.0050	0.025	0.050	ug/L
60-57-1	Dieldrin	0.025	U	0.0047	0.025	0.050	ug/L
72-55-9	4,4-DDE	0.025	U	0.0045	0.025	0.050	ug/L
72-20-8	Endrin	0.010	U	0.0043	0.010	0.050	ug/L
33213-65-9	Endosulfan II	0.025	U	0.0075	0.025	0.050	ug/L
72-54-8	4,4-DDD	0.025	U	0.0092	0.025	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.025	U	0.0035	0.025	0.050	ug/L
50-29-3	4,4-DDT	0.025	U	0.0044	0.025	0.050	ug/L
72-43-5	Methoxychlor	0.025	U	0.011	0.025	0.050	ug/L
53494-70-5	Endrin ketone	0.025	U	0.0097	0.025	0.050	ug/L
7421-93-4	Endrin aldehyde	0.025	U	0.0099	0.025	0.050	ug/L
5103-71-9	alpha-Chlordane	0.025	U	0.0060	0.025	0.050	ug/L
5103-74-2	gamma-Chlordane	0.025	U	0.0060	0.025	0.050	ug/L
8001-35-2	Toxaphene	0.50	U	0.15	0.50	1.00	ug/L
57-74-9	Chlordane	0.25	U	0.082	0.25	0.50	ug/L
2385-85-5	Mirex	0.025	U	0.0041	0.025	0.050	ug/L
SURROGATES							
2051-24-3	Decachlorobiphenyl	13.4		30 - 135		67%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.7		44 - 124		103%	SPK: 20

Report of Analysis

Client:	Weston Solutions		Date Collected:	01/28/25	
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169		Date Received:	01/29/25	
Client Sample ID:	TAPIAL2-MW03-012825-00-T3		SDG No.:	Q1211	
Lab Sample ID:	Q1211-02		Matrix:	WATER	
Analytical Method:	SW8081		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PESTICIDE Group1	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093939.D	1	01/30/25 09:25	01/31/25 14:15	PB166365

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	Q1211	OrderDate:	1/29/2025 10:10:00 AM
Client:	Weston Solutions	Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169
Contact:	Nathan Fretz	Location:	N31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1211-01	TAPHHA-MW01-01282 5-00-T4	Water			01/28/25			01/29/25
			Diesel Range Organics	8015D		01/30/25	01/30/25	
			Gasoline Range Organics	8015D			01/29/25	
			PESTICIDE Group1	8081B		01/30/25	01/31/25	
Q1211-02	TAPIAL2-MW03-0128 25-00-T3	Water			01/28/25			01/29/25
			Diesel Range Organics	8015D		01/30/25	01/30/25	
			Gasoline Range Organics	8015D			01/29/25	
			PESTICIDE Group1	8081B		01/30/25	01/31/25	



SAMPLE DATA

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPHHA-MW01-012825-00-T4	SDG No.:	Q1211
Lab Sample ID:	Q1211-01	Matrix:	Water
Analytical Method:	8015D DRO	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	1 mL
Extraction Type:		Test:	Diesel Range Organics
GPC Factor :	PH :	Injection Volume :	
Prep Method :	SW3510		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FE052162.D	1	01/30/25 08:33	01/30/25 16:07	PB166364

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
DRO	DRO	25.0	J	10.0	25.0	50.0	ug/L
SURROGATES							
16416-32-3	Tetracosane-d50	17.0		29 - 130		85%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPIAL2-MW03-012825-00-T3	SDG No.:	Q1211
Lab Sample ID:	Q1211-02	Matrix:	Water
Analytical Method:	8015D DRO	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	1 mL
Extraction Type:		Test:	Diesel Range Organics
GPC Factor :	PH :	Injection Volume :	
Prep Method :	SW3510		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FE052163.D	1	01/30/25 08:33	01/30/25 16:37	PB166364

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
DRO	DRO	28.0	J	10.0	25.0	50.0	ug/L
SURROGATES							
16416-32-3	Tetracosane-d50	15.3		29 - 130		76%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	Q1211	OrderDate:	1/29/2025 10:10:00 AM
Client:	Weston Solutions	Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169
Contact:	Nathan Fretz	Location:	N31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1211-01	TAPHHA-MW01-01282 5-00-T4	Water			01/28/25	01/30/25	01/30/25	01/29/25
			Diesel Range Organics	8015D				
			Gasoline Range Organics	8015D				
Q1211-02	TAPIAL2-MW03-0128 25-00-T3	Water			01/28/25	01/30/25	01/30/25	01/29/25
			Diesel Range Organics	8015D				
			Gasoline Range Organics	8015D				

Hit Summary Sheet SW-846

SDG No.: Q1211 **Order ID:** Q1211
Client: Weston Solutions **Project ID:** Ft Meade Tipton Airfield Parcel RI - PO 01

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID : TAPHHA-MW01-012825-00-T4									
Q1211-01	TAPHHA-MW01-012825-00-T4	Water	Aluminum	8.41	J	0.99	5.00	10.0	ug/L
Q1211-01	TAPHHA-MW01-012825-00-T4	Water	Arsenic	0.045	J	0.045	0.13	0.50	ug/L
Q1211-01	TAPHHA-MW01-012825-00-T4	Water	Barium	11.3		0.15	0.63	5.00	ug/L
Q1211-01	TAPHHA-MW01-012825-00-T4	Water	Beryllium	0.080	J	0.080	0.13	0.50	ug/L
Q1211-01	TAPHHA-MW01-012825-00-T4	Water	Calcium	13300		31.3	95.0	250	ug/L
Q1211-01	TAPHHA-MW01-012825-00-T4	Water	Cobalt	1.01		0.031	0.13	0.50	ug/L
Q1211-01	TAPHHA-MW01-012825-00-T4	Water	Copper	0.21	J	0.20	0.75	1.00	ug/L
Q1211-01	TAPHHA-MW01-012825-00-T4	Water	Iron	29.0		4.80	12.5	25.0	ug/L
Q1211-01	TAPHHA-MW01-012825-00-T4	Water	Lead	0.18	J	0.055	0.38	0.50	ug/L
Q1211-01	TAPHHA-MW01-012825-00-T4	Water	Magnesium	2790		13.3	95.0	250	ug/L
Q1211-01	TAPHHA-MW01-012825-00-T4	Water	Manganese	148		0.12	0.38	0.50	ug/L
Q1211-01	TAPHHA-MW01-012825-00-T4	Water	Nickel	0.56		0.090	0.13	0.50	ug/L
Q1211-01	TAPHHA-MW01-012825-00-T4	Water	Potassium	1350		23.1	95.0	250	ug/L
Q1211-01	TAPHHA-MW01-012825-00-T4	Water	Selenium	1.22	J	0.69	2.25	2.50	ug/L
Q1211-01	TAPHHA-MW01-012825-00-T4	Water	Sodium	5090		42.9	95.0	250	ug/L
Q1211-01	TAPHHA-MW01-012825-00-T4	Water	Zinc	2.32	J	0.28	0.75	2.50	ug/L
Q1211-01	TAPHHA-MW01-012825-00-T4	Water	Hardness, Total	44700		133	628	1650	ug/L
Client ID : TAPIAL2-MW03-012825-00-T3									
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Aluminum	93.0		0.99	5.00	10.0	ug/L
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Arsenic	2.26		0.045	0.13	0.50	ug/L
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Barium	20.4		0.15	0.63	5.00	ug/L
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Beryllium	0.080	J	0.080	0.13	0.50	ug/L
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Calcium	6660		31.3	95.0	250	ug/L
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Chromium	0.60	J	0.20	0.38	1.00	ug/L
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Cobalt	3.61		0.031	0.13	0.50	ug/L
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Copper	0.42	J	0.20	0.75	1.00	ug/L
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Iron	18500		4.80	12.5	25.0	ug/L
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Lead	0.17	J	0.055	0.38	0.50	ug/L
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Magnesium	1440		13.3	95.0	250	ug/L
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Manganese	541		0.12	0.38	0.50	ug/L
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Nickel	1.00		0.090	0.13	0.50	ug/L
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Mercury	0.090	J	0.081	0.16	0.20	ug/L
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Potassium	1250		23.1	95.0	250	ug/L
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Sodium	864		42.9	95.0	250	ug/L
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Vanadium	1.50	J	0.036	0.13	2.50	ug/L
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Zinc	2.02	J	0.28	0.75	2.50	ug/L
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	Hardness, Total	22600		133	628	1650	ug/L



SAMPLE DATA

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPHHA-MW01-012825-00-T4	SDG No.:	Q1211
Lab Sample ID:	Q1211-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	8.41	JN*	1	0.99	5.00	10.0	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7440-36-0	Antimony	0.13	U	1	0.055	0.13	1.00	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7440-38-2	Arsenic	0.045	JN	1	0.045	0.13	0.50	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7440-39-3	Barium	11.3		1	0.15	0.63	5.00	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7440-41-7	Beryllium	0.080	J	1	0.080	0.13	0.50	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7440-43-9	Cadmium	0.25	U	1	0.15	0.25	0.50	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7440-70-2	Calcium	13300	N	1	31.3	95.0	250	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7440-47-3	Chromium	0.38	U	1	0.20	0.38	1.00	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7440-48-4	Cobalt	1.01		1	0.031	0.13	0.50	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7440-50-8	Copper	0.21	J	1	0.20	0.75	1.00	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
Hardness	Hardness, Total	44700		1	133	628	1650	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7439-89-6	Iron	29.0	N	1	4.80	12.5	25.0	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7439-92-1	Lead	0.18	J	1	0.055	0.38	0.50	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7439-95-4	Magnesium	2790		1	13.3	95.0	250	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7439-96-5	Manganese	148		1	0.12	0.38	0.50	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7439-97-6	Mercury	0.16	UN	1	0.081	0.16	0.20	ug/L	02/06/25 08:55	02/06/25 12:13	SW7470A	
7440-02-0	Nickel	0.56		1	0.090	0.13	0.50	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7440-09-7	Potassium	1350	N	1	23.1	95.0	250	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7782-49-2	Selenium	1.22	J	1	0.69	2.25	2.50	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7440-22-4	Silver	0.25	UN	1	0.039	0.25	0.50	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7440-23-5	Sodium	5090		1	42.9	95.0	250	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7440-28-0	Thallium	0.25	U	1	0.043	0.25	0.50	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7440-62-2	Vanadium	0.13	U	1	0.036	0.13	2.50	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A
7440-66-6	Zinc	2.32	J	1	0.28	0.75	2.50	ug/L	01/30/25 10:30	02/06/25 18:31	SW6020	3010A

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	METALS-TAL			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPIAL2-MW03-012825-00-T3	SDG No.:	Q1211
Lab Sample ID:	Q1211-02	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	93.0	N*	1	0.99	5.00	10.0	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7440-36-0	Antimony	0.13	U	1	0.055	0.13	1.00	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7440-38-2	Arsenic	2.26	N	1	0.045	0.13	0.50	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7440-39-3	Barium	20.4		1	0.15	0.63	5.00	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7440-41-7	Beryllium	0.080	J	1	0.080	0.13	0.50	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7440-43-9	Cadmium	0.25	U	1	0.15	0.25	0.50	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7440-70-2	Calcium	6660	N	1	31.3	95.0	250	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7440-47-3	Chromium	0.60	J	1	0.20	0.38	1.00	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7440-48-4	Cobalt	3.61		1	0.031	0.13	0.50	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7440-50-8	Copper	0.42	J	1	0.20	0.75	1.00	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
Hardness	Hardness, Total	22600		1	133	628	1650	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7439-89-6	Iron	18500	N	1	4.80	12.5	25.0	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7439-92-1	Lead	0.17	J	1	0.055	0.38	0.50	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7439-95-4	Magnesium	1440		1	13.3	95.0	250	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7439-96-5	Manganese	541		1	0.12	0.38	0.50	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7439-97-6	Mercury	0.090	JN	1	0.081	0.16	0.20	ug/L	02/06/25 08:55	02/06/25 12:15	SW7470A	
7440-02-0	Nickel	1.00		1	0.090	0.13	0.50	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7440-09-7	Potassium	1250	N	1	23.1	95.0	250	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7782-49-2	Selenium	2.25	U	1	0.69	2.25	2.50	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7440-22-4	Silver	0.25	UN	1	0.039	0.25	0.50	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7440-23-5	Sodium	864		1	42.9	95.0	250	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7440-28-0	Thallium	0.25	U	1	0.043	0.25	0.50	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7440-62-2	Vanadium	1.50	J	1	0.036	0.13	2.50	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A
7440-66-6	Zinc	2.02	J	1	0.28	0.75	2.50	ug/L	01/30/25 10:30	02/06/25 18:34	SW6020	3010A

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	METALS-TAL			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

LAB CHRONICLE

OrderID:	Q1211	OrderDate:	1/29/2025 10:10:00 AM
Client:	Weston Solutions	Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169
Contact:	Nathan Fretz	Location:	N31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1211-01	TAPHHA-MW01-01282 5-00-T4	Water			01/28/25			01/29/25
			Mercury	7470A		02/06/25	02/06/25	
			Metals ICP-TAL	6020B		01/30/25	02/06/25	
Q1211-02	TAPIAL2-MW03-0128 25-00-T3	Water			01/28/25			01/29/25
			Mercury	7470A		02/06/25	02/06/25	
			Metals ICP-TAL	6020B		01/30/25	02/06/25	



SAMPLE DATA

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25 12:00
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPHHA-MW01-012825-00-T4	SDG No.:	Q1211
Lab Sample ID:	Q1211-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Ammonia as N	0.080	U	1	0.045	0.080	0.10	mg/L	02/03/25 08:50	02/03/25 13:39	SM 4500-NH3 B plus G-11
Bromide	1.00	U	1	0.034	1.00	2.00	mg/L		01/29/25 12:41	9056A
Chloride	17.8	OR	1	0.011	0.30	0.60	mg/L		01/29/25 12:41	9056A
Fluoride	0.070	J	1	0.057	0.20	0.40	mg/L		01/29/25 12:41	9056A
Nitrite	0.30	U	1	0.011	0.30	0.60	mg/L		01/29/25 12:41	9056A
Nitrate	0.69		1	0.0034	0.25	0.50	mg/L		01/29/25 12:41	9056A
Sulfate	25.3		1	0.032	1.50	3.00	mg/L		01/29/25 12:41	9056A
Dissolved Hexavalent Chromium	0.0050	U	1	0.0030	0.0050	0.010	mg/L		01/29/25 14:45	7196A
Oil and Grease	2.00	U	1	0.40	2.00	5.00	mg/L		02/01/25 11:30	1664A
TOC	0.91	J	1	0.19	0.50	1.00	mg/L		01/29/25 14:30	9060A

Comments: _____

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25 12:00
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPHHA-MW01-012825-00-T4DL	SDG No.:	Q1211
Lab Sample ID:	Q1211-01DL	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Chloride	16.7	D	5	0.055	1.50	3.00	mg/L		01/29/25 14:07	9056A

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits

Report of Analysis

Client:	Weston Solutions	Date Collected:	01/28/25 14:55
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	01/29/25
Client Sample ID:	TAPIAL2-MW03-012825-00-T3	SDG No.:	Q1211
Lab Sample ID:	Q1211-02	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Ammonia as N	0.40		1	0.045	0.080	0.10	mg/L	02/03/25 08:50	02/03/25 13:50	SM 4500-NH3 B plus G-11
Bromide	1.00	U	1	0.034	1.00	2.00	mg/L		01/29/25 13:46	9056A
Chloride	3.10		1	0.011	0.30	0.60	mg/L		01/29/25 13:46	9056A
Fluoride	0.076	J	1	0.057	0.20	0.40	mg/L		01/29/25 13:46	9056A
Nitrite	0.30	U	1	0.011	0.30	0.60	mg/L		01/29/25 13:46	9056A
Nitrate	0.25	U	1	0.0034	0.25	0.50	mg/L		01/29/25 13:46	9056A
Sulfate	30.7		1	0.032	1.50	3.00	mg/L		01/29/25 13:46	9056A
Dissolved Hexavalent Chromium	0.0050	U	1	0.0030	0.0050	0.010	mg/L		01/29/25 14:49	7196A
Oil and Grease	2.00	U	1	0.40	2.00	5.00	mg/L		02/01/25 11:30	1664A
TOC	7.30		1	0.19	0.50	1.00	mg/L		01/29/25 15:47	9060A

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

LAB CHRONICLE

OrderID:	Q1211	OrderDate:	1/29/2025 10:10:00 AM
Client:	Weston Solutions	Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169
Contact:	Nathan Fretz	Location:	N31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1211-01	TAPHHA-MW01-01282 5-00-T4	WATER			01/28/25 12:00			01/29/25
			Ammonia	SM4500-NH3		02/03/25	02/03/25 13:39	
			Anions Group5	9056A			01/29/25 12:41	
			Hexavalent Chromium	7196A			01/29/25 14:45	
			Oil and Grease	1664A			02/01/25 11:30	
			TOC	9060A			01/29/25 14:30	
Q1211-01DL	TAPHHA-MW01-01282 5-00-T4DL	WATER			01/28/25 12:00			01/29/25
			Anions Group5	9056A			01/29/25 14:07	
Q1211-02	TAPIAL2-MW03-0128 25-00-T3	WATER			01/28/25 14:55			01/29/25
			Ammonia	SM4500-NH3		02/03/25	02/03/25 13:50	
			Anions Group5	9056A			01/29/25 13:46	
			Hexavalent Chromium	7196A			01/29/25 14:49	
			Oil and Grease	1664A			02/01/25 11:30	
			TOC	9060A			01/29/25 15:47	



SHIPPING DOCUMENTS

Q1211

12

12.1



Weston COC ID
Weston_20250128_1605

Chain of Custody Record/Lab Work Request

Page	1	of	1
------	---	----	---

Client:	Weston Solutions, Inc.		
Project Manager:	David Sembrot		
Street Address:	1400 Weston Way	City:	West Chester
Phone:	610-314-5456	ST, ZIP:	PA, 19038
e-mail:	david.sembrot@westonsolutions.com		
Sampled By:	Cheyenne Harrington		


Project Name:	Fort Meade RI	Project POC:	Nathan Fretz
PO Number	0111169	Phone:	484-524-5665
W.O. #:		POC e-mail:	nathan.fretz@westonsolutions.com
Lab:	CHEMTECH	Lab POC:	Yazmeen Gomez
TAT (days):	21	Lab Phone:	908-728-3144
Lab Address:	284 Sheffield Street Mountainside, NJ 07092		

Matrix Codes
SS - Soil
SE - Sediment
SO - Solid
SL - Sludge
GW - Groundwater
W - Water
SB - Soil Boring
A - Air
DS - Drum Solids
DL - Drum Liquids
L - EP/TCLP Leachate
WI - Wipe
X - Other
F - Fish

Lab Use Only		
Temperature of cooler when received (°C)		
COC Tape was present and unbroken on outer package?	Y	N
Samples received in good condition?	Y	N
Labels indicate properly preserved?	Y	N
Received within holding times?	Y	N
Discrepancies between sample labels and COC record?	Y	N

Analyses Requested:	DRO by EPA 8015D	Pesticides by EPA 8081B	SVOCs by EPA 8270E	O&G by EPA 1664A	Hardness by EPA 200	Anions by EPA 9056A	TOC by EPA 9060A/Lloyd Kahn	GRO by EPA 8015D	VOCs by EPA 8260D	Hex Cr by EPA	Ammonia by SM4500-NH3 B P	Metals w/ Hg by EPA 6020B/17470A
Container Type:	Amber	Amber	Amber	Glass	Plastic	Plastic	Vial	Vial	Vial	Plastic	Plastic	Plastic
Container Size:	1 L	1 L	1 L	1 L	1 L	1 L	40 mL	40 mL	40 mL	500 mL	500 mL	500 mL
Preservative:	Ice to 0-6	Ice to 0-6	Ice to 0-6	H2SO4 to <	HNO3 to pH	Ice to 0-6	H2SO4 to <	HCL to PH	HCL to Ph	Ammonium	H2SO4: Ice	HNO3 to pH

#	Sample ID	G/C	Matrix	# Cont	MS/MSD	Date Collected	Time Collected	X	X	X	X	X	X	X	X	X	X	X	Special Instructions/Comments
1	TAPHHA-MW01-012825-00-T4	g	GW	19	no	1/28/2025	12:00	X	X	X	X	X	X	X	X	X	X	X	
2	TAPIAL2-MW03-012825-00-T3	g	GW	19	no	1/28/2025	14:55	X	X	X	X	X	X	X	X	X	X	X	
3	TAP-TB-02-012825-T4	g	W	2	no	1/28/2025	12:05								X				
4																			
5																			
6																			
7																			
8																			
9																			
10																			
11																			
12																			

Shipping Airbill Number: 77173295 4230, 77173295 4240						Cooler Number: 1/2 of 2	
Relinquished By		Date	Time	Received By	Date	Time	Additional Comments
1.)		1/28/25	1710		1-29-25	10:00	QSM 6.0 Compliant
2.)							Deliverable Requirements: DoD Level IV report, EnviroData EDD, and ERIS-compatible EDD
3.)							

2.1", 2.3"

73 of 75

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1211 WEST04

Order Date : 1/29/2025 10:10:00 AM

Project Mgr :

Client Name : Weston Solutions

Project Name : Ft Meade Tipton Airfield Pa

Report Type : Level 4

Client Contact : Nathan Fretz

Receive DateTime : 1/29/2025 10:00:00 AM

EDD Type : SEDD 2A

Invoice Name : Weston Solutions

Purchase Order :

Hard Copy Date :

Invoice Contact : Nathan Fretz

Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1211-01	TPHHA-MW01-012825-00-T4 TAPHHA-MW01-012825-00-T4	Water	01/28/2025	12:00	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q1211-02	TAPIAL2-MW03-012825-00-T3	Water	01/28/2025	14:55	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q1211-03	TAP-TB-02-012825-T4	Water	01/28/2025	12:05	VOC-TCLVOA-10		8260D	10 Bus. Days	

Relinquished By :

Date / Time : 1-29-25 11:25

Received By :

Date / Time :

Storage Area : VOA Refridgerator Room